



Date: -08th June, 2017

Corrigendum
For
Tender for
Instruments for Shoulder Arthroscopy and Open
Shoulder Surgery for the Department of Orthopaedics

NIT Issue Date	: 03 rd April, 2017
NIT No.	: Admn/Tender/10/2017-AIIMS.JDH
Pre-Bid Meeting	: 13 th April, 2017 at 03:45 PM
Earlier Last Date of Submission	: 15 th June, 2017 at 03:00 PM
Extended Last Date of Submission	: 21 st June, 2017 at 03:00 PM
Bid opening	: 22 nd June, 2017 at 03:30 P.M

The following revised and additional specification will be added:-

- 1. Page No. 12, In Technical Specification, Shoulder Arthroscopy Instruments, Shoulder Limb Positioner and Accessories, Point No. 01:**
For
Suture Manipulator with atraumatic tip to spread parallel strands of suture with its jaw to create a closed loop to allow suture slide freely during extraction
Read
Suture Manipulator with atraumatic tip
- 2. Page No. 12, Point No. 07:**
For
Tissue penetrator cum suture retriever in cigar handle with small sharp penetrator curved tip so that it can be used to grasp the labrum at 6'o clock position
Read
Tissue penetrator cum suture retriever in cigar handle / loop handle
- 3. Page No. 12, Point No. 08:**
For
Multifunctional suture passer with provision of only 1 needle, which can hold the cuff & at the same time pass the suture & retrieve it in one step
Read
Multifunctional suture passer which can hold the cuff & at the same time pass the suture & retrieve it in one step

4. Page No. 12, Point No. 10:

For

Switching stick 4.3mm, non cannulated, 13" long

Read

Switching stick 4-5mm, non cannulated

5. Page No. 13, In technical Specification, Reusable cannulated obturator, Sub Line No. 01 to 03:

For

8.5mm cannula for easy cannula insertion

7mm cannula for easy cannula insertion

5.5mm cannula for easy cannula insertion

Read

7.5-8.5mm cannula for easy cannula insertion

6-7mm cannula for easy cannula insertion

5-6mm cannula for easy cannula insertion