



Date: - 19th August, 2019

Corrigendum
For
Adjustable LED Surgical Head Light with
attachment for the Department of Surgical
Gastroenterology

NIT Issue Date	: 17 th June, 2019
NIT No.	: Admn/Tender/129/2018-AIIMS.JDH
Pre-Bid Meeting	: 26 th June, 2019 at 03:15 PM
Earlier Date of Submission	: 13 th August, 2019 at 03:00 PM
Extended Date of Submission	: 29 th August, 2019 at 03:00 PM
Bid opening	: 30 th August, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page 10, In technical specification, point b)**
For:
Should have a maximum luminous Intensity of around 2,00,000 lux or more
Read:
Should have a luminous Intensity of around 70,000 lux or more
- 2. Page 10, In technical specification, point c)**
For:
Should have color temperature of 5000K
Read:
Should have color temperature of 4500-5000K
- 3. Page 10, In technical specification, point g)**
For:
Should have integrated battery status indicators
Read:
Should have integrated battery status indicators/Audible Alarm
- 4. Page 10, In technical specification, point j)**
For:
Should be able to focus down to about 65mm.
Read:
Should be able to focus down to about 65 -70mm.
- 5. Page 10, In technical specification, point l)**
For:
Should have an extra spare battery
Read:
Should have 2 extra spare battery
- 6. All items must be US-FDA or European CE (with four digit notified body number) approved.**