



Date: - 29th August, 2020

Corrigendum
for
Flexible Video Cystonephroscope with Chip on the
tip for Urological Procedure with high Definition
System with Recording and Trolley for the
Department of Urology

NIT Issue Date	: 20 th July, 2020
NIT No.	: Admn/Tender/135-3/2018-AIIMS.JDH
Pre-Bid Meeting	: 30 th July 2020 at 04:15 PM
Earlier Last Date of Submission	: 31 st August, 2020 at 03:00 PM
Extended Last Date of Submission	: 15 th September, 2020 at 03:00 PM
Bid opening	: 16 th September, 2020 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page No.10, In technical specification, point No. 11):**
For
300 Watt xenon or LED light equivalent to 300 watt xenon: 1 No a. A Powerful Should be 300W LED/Xenon Light Source b. Automatically adjust light intensity to achieve ideal illumination. c. Xenon Light source should have minimum 500 Hours bulb life or LED greater than 10000 hours.
Read
300 watt Xenon or LED light source should have 30000 lamp hours and color temperature of 6500 K.
- 2. Page No.10, In technical specification, point No. 12:**
For
• 26" Full HD Medical Grade Monitor: 1 No's
Read
• 26-27" Full HD Medical Grade Monitor: 1 No's
- 3. Page No.11, In technical specification, point No. 13) g:**
For
• Recording system should be supporting 2D and 3D full HD recording
Read
• Recording system should be supporting 2D and Gaute full HD recording.
- 4. Page No.11, In technical specification, point No. 14:**
For •
Flexible Videocysto nephroscope - 2 No's

Read:

- Flexible Videocysto nephroscope - 1 No's

5. Page No. 11, In technical specification, point No. 14) r:**For:**

- Suction/Irrigation facility should be available

Read:

- Deleted

6. Page No.12, In technical specification, pointNo. 16:**For:****Accessories**

- Compatible flexible biopsy forceps - 4 No's
- Compatible flexible DJ stent removal forceps - 5 No's
- Nitinol tipples stone extractor 1.5 Fr. and length 115 cm - 5 No's
- Nitinol tipples stone extractor 2 Fr. and length 115 cm - 5 No's
- Nitinol multi-wire (12-16 wire basket) geometry stone extractor (partial closer) 1.7 Fr. and length 115 cm - 5 No's
- Nitinol multi-wire (12-16 wire basket) geom~try stone extractor (partial closer) 2.4 Fr. and length 115 cm - 5 No's
- Ngage nitinol stone extractor 1.7 Fr. and 115 cm. - 5 No's
- Ngage nitinol stone extractor 2.2 Fr. and 115 cm. - 5 No's
- Nitinol core guide wire with hydrophilic coating and should have ability being straight at one end and angled at other end 0.032" and length 150 cm - 20 No's
- Compatible XENON bulbs for light source - 2 No's

Read:**Accessories**

- Compatible flexible biopsy forceps - 2 No's
- Compatible flexible DJ stent removal forceps - 2 No's
- Nitinol tipples stone extractor 1.5 Fr. and length 115 cm -1 No's
- Nitinol tipples stone extractor 2 Fr. and length 115 cm - 2 No's
- Nitinol multi-wire (12-16 wire basket) geometry stone extractor (partial closer) 1.7 Fr. and length 115 cm -1 No's
- Nitinol multi-wire (12-16 wire basket) geometry stone extractor (partial closer) 2.4 Fr. and length 115 cm - 2 No's
- Ngage nitinol stone extractor 1.7 Fr. and 115 cm. -1 No's
- Ngage nitinol stone extractor 2.2 Fr. and 115 cm. - 2 No's
- Nitinol core guide wire with hydrophilic coating and should have ability being straight at one end and angled at other end 0.032" and length 150 cm - 10 No's
- Compatible XENON bulbs for light source - 2 No's