



Date: - 19<sup>th</sup> August, 2019

Corrigendum  
For  
Operating Microscope for Spine Surgery for the  
Department of Orthopaedics

NIT Issue Date	: 01 <sup>st</sup> July, 2019
NIT No.	: Admn/Tender/143-2/2018-AIIMS.JDH
Pre-Bid Meeting	: 11 <sup>th</sup> July, 2019 at 04:00 PM
Earlier Date of Submission	: 13 <sup>th</sup> August, 2019 at 03:00 PM
Extended Date of Submission	: 29 <sup>th</sup> August, 2019 at 03:00 PM
Bid opening	: 30 <sup>th</sup> August, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page No. 10, Technical Specification, S. No.7:**  
**For**  
Material - Stainless Steel (Instruments)  
**Read**  
Deleted
- 2. Page No. 10, Technical Specification, S. No.8:**  
**For**  
Grade - ISO 5832-1  
**Read**  
Deleted
- 3. Page No. 10, Technical Specification, AUTODRAPE,:**  
**For**  
Preferably System Should be capable of Draping the microscope by push of a button.  
**Read**  
System should be capable of Draping the microscope by push of a button.
- 4. Page No. 10, Technical Specification, FLOOR STAND, Point No.4:**  
**For**  
System Should Preferably have facility for Auto Air evacuation system for drapes (Auto drapes). System should have overhead display for showing important parameters.  
**Read**  
System should have facility for auto air evacuation system for drapes (Auto drapes). System should have overhead display for showing important parameters.