



Date: - 23rd July, 2020

Corrigendum
for
Ultrasonic Surgical Aspirator for the Department of
Surgical Gastroenterology

NIT Issue Date	: 23 rd May, 2020
NIT No.	: Admn/Tender/15/2020-AIIMS.JDH
Pre-Bid Meeting	: 03 rd June, 2020 at 03:45 PM
Earlier Last Date of Submission	: 30 th July, 2020 at 03:00 PM
Extended Last Date of Submission	: 05 th August, 2020 at 03:00 PM
Bid opening	: 06 th August, 2020 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Point No. '2':

For

The system should be based on Magneto restrictive and Piezoelectric Technology with 23 — 40 KHz frequency.

Read as:

The system should be based on Magneto restrictive or Piezoelectric Technology with 23 - 40 KHz frequency.

2. Page No. 10, Point No. '15', bullet no. 1 :

For

23 and 36 KHz Universal Straight Handpiece -1 No Each

Read as:

Hand pieces: Each having frequencies range from 23-25 KHz and 34-36 KHz (Universal Straight Handpiece) -1 No. Each

3. Page No. 10, point 17 :

For

The control panel should have adjustable viewing angle for better visibility in the O.R. and must be equipped with quite pumps.

Read As:

The control panel should have adjustable viewing angle for better visibility in the O.R. (optional) and must be equipped with quite pumps.

4. Page No. 11, Addition of Point 20":

The ultrasonic surgical aspirator system should have laparoscopic mode that provides no suction to the handpiece until the vibration pedal is activated which helps to maintain the gas during the procedure.