



Date:- 10th March 2022

Corrigendum  
for  
Fully Random Access Chemiluminescence (Reagent  
rental basis)

NIT Issue Date	: 01 <sup>st</sup> February, 2022
NIT No.	: Admn/Tender/16/2021-AIIMS.JDH
Pre-Bid Meeting	: 15 <sup>th</sup> February, 2022 at 03:00 PM
Earlier Last Date of Submission	: 08 <sup>th</sup> March, 2022 at 03:00 PM
Extended Last Date of Submission	: 28 <sup>th</sup> March, 2022 at 03:00 PM
Bid opening	: 29 <sup>th</sup> March, 2022 at 03:15 P.M

**The following revised and additional specification will be added: -**

**1. Page No.8, Point no 8**

**For**

The instrument should have a facility of lot calibration, auto loading & unloading of reagents while instrument is in running mode

**Read**

The instrument should have a facility of lot calibration and preferably auto loading & unloading of reagents while instrument is in running mode