



Date: - 22nd July, 2020

Corrigendum for Renal Transplant Instruments for the Department of Urology

NIT Issue Date	: 20 th May, 2020
NIT No.	: Admn/Tender/17/2020-AIIMS.JDH
Pre-Bid Meeting	: 03 rd June, 2020 at 04:15 PM
Earlier Last Date of Submission	: 30 th July, 2020 at 03:00 PM
Extended Last Date of Submission	: 05 th August, 2020 at 03:00 PM
Bid opening	: 06 th August, 2020 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 20, In technical specification, In Retractor set Toggle Button No. 1:

For

BALFOUR ABD RETRACTORW/BV609BV624TOT

Read

Balfour blades W/lips 70 mm x 76 mm (2 3/4" x 6 7/8")

Balfour blades W/lips 83 mm x 76 mm (3 1/4" x 3")

Wide Balfour blades 83 mm x 57 mm (3 1/4" x 1/4")

Balfour blades 83 mm x 76 mm (3 1/4" x 3")

2. Page No. 20, In technical specification, In Retractor set Toggle Button No. 11:

For

OMNI-TRACT SURGICAL RETRACTION SYSTEM WITH BLADES

Read

Instruments	Quantity (Nos.) 18-4301C UROLOGY LOBSTER	Quantity (Nos.) 18-4301C UROLOGY Expanded LOBSTER
Vertical Post With Table Rail Clamp & Coupling Clamp	1	1
Lobster TM Extension Arm	-	1
Lobster TM Frame	1	1
Fence Retractor 4" x 6" (10x 15cm)	1	1
Deaver Retractor 1 1/2 x 6" (3.8 x 15.2cm)	1	1

Deaver Retractor 2 3/4" x 4 1/2" (7.0 x 11.4cm)	1	1
Malleable Retractor Perforated 2" x 6 1/2" (5.0 x 16.5cm)	1	1
Malleable Retractor, Perforated 3" x 6 1/2" (7.6 x 16.5cm)	1	1
Malleable Bladder Retractor 3" x 6" (7.6 x 15.2cm)	1	1
Harrington 'Sweetheart' Retractor 2" x 5" (5 x 12.5cm)	1	1
Harrington 'Sweetheart' Retractor 3" x 7" (7.6 x 17.8cm)	1	1
Splanchnic Retractor 2" x 4 1/2" (5.0 x 11.4cm)	1	1
Mayo Retractor 2" x 2" (5.0 x 5.0 cm)	2	2
Mayo Retractor 2 3/4" x 2" (7.0 x 5.0cm)	2	2
Retractor Clamp	6	8

3. Page No. 20, In technical specification, In Retractor set add Toggle Button No. 14:

ADD:

FINOCHIETTO BURFORD RETRACTOR (aluminium):

Single blade with two pairs of detachable blades, length 300 mm, 12"; maximum spread 190-220 mm
2 Nos.

1. Blades: 45 x 64 mm - 2 pairs
2. Blades: 75 x 64 mm - 2 pairs

4. Page No. 22, In technical specification, In Maintenance of Surgical Instruments Toggle Button No. 1:

For

STERILIT I OIL SPRAY

Read

Lubricant OIL SPRAY

5. Page No. 22, In technical specification:

Point Added

Other:

1. Each instrument must have the **LASER engraved** catalogue number and logo / manufacturer name on it.
2. Demonstration of the instrument's to be provided by the firm.
3. Bidder should quote all items. 80-90 % items to be quoted from single manufacturer, remaining 10-20% can be from different vendor of same quality. + - 10% variation in dimension of the instrument is acceptable.
4. All instrument must be made of corrosion free surgical grade stainless steel (Stainless steel alloy of 350-450 grade)
5. All Instruments should be USFDA/ Genuine European CE four digits certified.