



Date: - 27<sup>th</sup> November, 2017

Corrigendum  
For  
Tender for Microsurgery Drill System for the  
Department of Neurosurgery.

<b>NIT Issue Date</b>	<b>: 06<sup>th</sup> October, 2017</b>
<b>NIT No.</b>	<b>: Admn/Tender/188/2017-AIIMS.JDH</b>
<b>Pre-Bid Meeting</b>	<b>: 16<sup>th</sup> October, 2017 at 03:00 PM</b>
<b>Earlier Last Date of Submission</b>	<b>: 04<sup>th</sup> December, 2017 at 03:00 PM</b>
<b>Extended Last Date of Submission</b>	<b>: 18<sup>th</sup> December, 2017 at 03:00 PM</b>
<b>Bid opening</b>	<b>: 19<sup>th</sup> December, 2017 at 03:15 P.M</b>

**The following revised and additional specification will be added:-**

**1. Page 9, point 1**

**For**

High Speed Electrical Drill system with Variable speed setting from 10,000 to  $\geq 80,000$  rpm..

**Read**

High Speed Electrical Drill system with Variable speed setting with maximum speed more than 60000 rpm.

**Page 9, two more points to be added after point 1.**

- Minimum torque should be 40 mNm.
- Drill should be light-weight (less than 150 gms) and ergonomically designed for long duration surgeries.

**2. Page 10, Heading II. Accessories, A. Perforator Set, point b**

**For**

Reusable cranial perforators for adult and pediatric sizes (10 each) to be supplied.

**Read**

Disposable cranial perforators for adult and pediatric sizes (10 each) to be supplied

**3. Page 10, Heading II. Accessories, B. Craniotomy Set, point c, item iii**

**For**

Large: 05 (for c)

**Read**

Deleted

**4. Page 10, Heading II. Accessories, Heading of C.**

**For**

Osteotomy saw set

**Read**

Osteotomy Reciprocating saw Hand piece

**5. Page 10, Heading II. Accessories, subheading C, point b.**

**For**

It should also have tools to harvest bone grafts from ribs or iliac crest.

**Read**

Delete

**6. Page 10, Heading II. Accessories, subheading C, point c**

**For**

Two set of necessary blades to be supplied.

**Read**

Two saw blades each of short and long sizes (Total 04 in numbers) to be supplied

**7. Page 10, Heading II. Accessories, subheading E, point b**

**For**

Should have angled long (approx. 14 to 16 cm +/-1) and short (approx. 10 to 12 cm +/-1) attachments of less than 5mm outer thickness for use in trans-sphenoidal, trans-oral, micro-spinal and endoscopic surgeries (preferably of variable length) – one each,

- With corresponding cutting burrs of 1mm, 2mm and 4mm heads – 5 each
- With corresponding diamond burrs of 1mm, 2mm and 4mm – 5 each

**Read**

Should have angled long (approx. 14 to 16 cm +/-1) and short (approx. 10 to 12 cm +/-1) attachments of less than 5mm +/- 1mm outer thickness for use in trans-sphenoidal, trans-oral, micro-spinal and endoscopic surgeries (preferably of variable length) – one each,

- With corresponding cutting burrs of 2 mm,3 mm and 4mm heads – 5 each
- With corresponding diamond burrs of 2mm, 3mm and 4mm – 5 each”

**8. Page 10, Heading III. Other essentials: 4 points to be added**

- a. Hand piece lubrication and cleaning solution- 6 in number
- b. Irrigation Tubing- 20 in number
- c. Imported cart to be supplied with the system.
- d. Demonstration of the product (part/full) may be demanded before approval.