



Date: - 03rd August 2016

Corrigendum
for
OT Light System

NIT Issue Date	:	17 th June 2016.
NIT No.	:	Admn/Tender/19/2016-AIIMS.JDH
Pre Bid Meeting held on	:	27 th June, 2016 at 03:00 PM
Last Date of Submission	:	18 th July, 2016 at 03:00 PM
Extended Last Date of Submission	:	04 th August, 2016 at 03:00 PM
Revised Last Date of Submission	:	23 rd August, 2016 at 03:00 PM

1. The following revised and additional specification will be added:-

1. Page No.9, Technical Specification, Point No. 01:

For

The Lights should have LED light engines in which the mixing of the various LED lights should take place inside the engines itself which should prevent the casting of color shadows.

Read

Should have single color white LED's or Light engines with mixing of various LED lights, to reduce color fringing and should prevent casting of color shadows.

2. Page No.9, Technical Specification, after point no. 04:

Added Point No. 4.A:

One Dome should be Camera ready for accepting HD Camera in future.

3. Page No.9, Technical Specification, Point No. 05:

For

Intensity at 1-meter distance 1,50,000 to 1,60,000 lux for major dome and not less than 1,10,000 to 1,30,000 lux for satellite dome.

Read

Intensity at 1-meter distance 1,50,000 to 1,60,000 lux for major dome and not less than 1,10,000 for satellite dome.

4. Page No.9, Technical Specification, Point No. 06:

For

Should have variable Colour Temperature: 3500-5500 K.

Read

Should have color temperature: 3500-5000 K.

5. Page No.9, Technical Specification, Point No. 11:

For

Illuminated field diameter should be approx. 20-30 cms.

Read

Illuminated field diameter should be approx. 25-30 cms and adjustable.

**Administrative Officer
AIIMS, Jodhpur**