



ALL INDIA INSTITUTE OF MEDICAL SCIENCES,  
JODHPUR  
\*\*\*\*\*

Dated: 10<sup>th</sup> December, 2024

**CORRIGENDUM**

**Rate Contract for Supply of Syringes, Needles, Angiocatheter and Dialysis Consumables**

**At**

**AIIMS, Jodhpur**

Tender No.	:	PROC-2/RC/12/2024-AIIMS.JDH
NIT Issue Date	:	27 <sup>th</sup> September, 2024
Last Date of Submission	:	12 <sup>th</sup> December, 2024 up to 03:00 PM
Revised last date of submission	:	<b>19<sup>th</sup> December, 2024 up to 03:00 PM</b>

**Page No.: 14 > Sample / Demonstration:**

**For**

- The bidders are required to submit samples of the items for which they have quoted (without indicating price, clear marking of firm / agency name in each of item / tender ref. number) **on or before last date of submission of the bids, failing which their bids/offer shall be rejected** and in case all the expenses will be borne by the tenderer.
- The samples are required to be submitted at **Central Store, Ground Floor, Near Amrit Pharmacy, IPD Building, AIIMS Jodhpur** in original packing, duly labelled (printed) and sealed having all relevant details such as manufacturing date, expiry date, batch number etc.
- The firm / vendor will have to submit samples of all such items for which they have participate in the tender. Bid without accompanying samples in the prescribed quantities / numbers will be summarily rejected. Every sample must have super scribed with the name of the firm, tender serial number in a separate envelope duly sealed, signed and stamped by the tenderer.

**Read As:**

- The bidders are required to submit samples of the items for which they have quoted (without indicating price, clear marking of firm / agency name in each of item / tender ref. number) **as and when required by the Institute, failing which their bids/offer shall be rejected** and in case all the expenses will be borne by the tenderer.
- The samples are required to be submitted at **Central Store, Ground Floor, Near Amrit Pharmacy, IPD Building, AIIMS Jodhpur** in original packing, duly labelled (printed) and sealed having all relevant details such as manufacturing date, expiry date, batch number etc.
- The firm / vendor will have to submit samples of all such items for which they have participate in the tender. Every sample must have super scribed with the name of the firm, tender serial number in a separate envelope duly sealed, signed and stamped by the tenderer.



**ALL INDIA INSTITUTE OF MEDICAL SCIENCES,  
JODHPUR**

\*\*\*\*\*

Page No.: 29 > Annexure – XIII > List of Items >

Amendments are as under:

T.R. No.	FOR:			READ AS:		
	Item Name	Specification	Size	Item Name	Specification	Size
259	Double lumen Tunneled Dialysis Palindrome catheters	Should have – Two lumens of Double – D design with symmetrical tip • Should have Squirrel Z tip design with laser cut Side Slots • Should have- Carbothane material • Should have- size 14.5 Fr with different lengths. • Should have- Vessel dilator with Flowguard to avoid blood spillage or air embolism • Should have- USFDA / European CE/ Approved / Equivalent Indian Standard Approved	14.5 fr	Double lumen Tunneled Dialysis catheters	Should have – Two lumens of Double – D design with symmetrical tip • Should have Squirrel Z tip design with laser cut Side Slots • Should have- Carbothane material • Should have- Vessel dilator with Flowguard to avoid blood spillage or air embolism • Should have- USFDA / European CE/ Approved / Equivalent Indian Standard Approved	14.5 fr x 19 cm
262	Double Lumen Tunneled Dialysis Palindrome catheters Si Ion & Heparin Coated(adult)	Should have- Two lumens of DoubleDesign with symmetrical tip • Should have Spiural Z tip design with Laser Cut side slots • Should have Heparin coating inside • Should have- Carbothane material • Should have- sizes 14.5 Fr with different lengths. • Should have- Vessel dilator with Flowguard to avoid blood spillage or air embolism USFDA / European CE/ Approved / Equivalent Indian Standard Approved		Double Lumen Tunneled Dialysis catheters Silver Ion & Heparin Coated(adult)	Should have- Two lumens of DoubleDesign with symmetrical tip • Should have Spiural Z tip design with Laser Cut side slots • Should have Heparin coating inside • Should have- Carbothane material • Should have- Vessel dilator with Flowguard to avoid blood spillage or air embolism USFDA / European CE/ Approved / Equivalent Indian Standard Approved	14.5 fr. X 19 cm
333	Double lumen HD Catheter	Should have two lumens of Double-D design Should have laser cut side slots to avoid positional occlusion Should have clear silicon extensions tubings in both venous and arterial port Should have 1-2 dilators in the pack Should have straight or curved extensions kits Flexible radiopaque polyurethane material with J end Guide wire. y shaped Double entry introducer needle with special syringe for introduction of Guide wire. Vessel dilator with arterial and venous port caps. Syringe 5-	12 fr*16 cm	Double lumen HD Catheter	<b>Should have two lumens of Double-D design Should have laser cut side slots to avoid positional occlusion Should have clear silicon extensions tubings in both venous and arterial port Should have 1-2 dilators in the pack Should have straight or curved extensions kits, Flexible radiopaque polyurethane material with J end Guide wire. 18 G introducer needle with special syringe for introduction of Guide wire. Vessel dilator with arterial and venous port caps. Syringe 5-10 cc LL. Sterilized. USFDA / European CE/</b>	11.5 fr*16 cm



**ALL INDIA INSTITUTE OF MEDICAL SCIENCES,  
JODHPUR**

\*\*\*\*\*

FOR:				READ AS:		
T.R. No.	Item Name	Specification	Size	Item Name	Specification	Size
		10 cc LL. Sterilized. USFDA / European CE/ Approved / Equivalent Indian Standard Approved			Approved / Equivalent Indian Standard Approved	

**Note:**

1. Above mentioned modifications will not reflect in BoQ, therefore, Bidders are advised to quoted price in front of old specification. **However, the modified specifications will be considered for evaluation as per the details mentioned above.**



# ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

\*\*\*\*\*

**Annexure – XIII > List of Items > Page No.: 44 >**

**Following items are added in existing list:**

S. No.	Item Name	Specification	Size
609.	Double lumen Tunneled Dialysis catheters	Should have – Two lumens of Double – D design with symmetrical tip • Should have Squiral Z tip design with laser cut Side Slots • Should have- Carbothane material • Should have- Vessel dilator with Flowguard to avoid blood spillage or air embolism • Should have- USFDA / European CE/ Approved / Equivalent Indian Standard Approved	14.5 fr x 23 cm
610.	Double lumen Tunneled Dialysis catheters	Should have – Two lumens of Double – D design with symmetrical tip • Should have Squiral Z tip design with laser cut Side Slots • Should have- Carbothane material • Should have- Vessel dilator with Flowguard to avoid blood spillage or air embolism • Should have- USFDA / European CE/ Approved / Equivalent Indian Standard Approved	14.5 fr x 28 cm
611.	Double Lumen Tunneled Dialysis catheters Silver Ion & Heparin Coated(adult)	Should have- Two lumens of DoubleDesign with symmetrical tip • Should have Spiural Z tip design with Laser Cut side slots • Should have Heparin coating inside • Should have- Carbothane material • Should have- Vessel dilator with Flowguard to avoid blood spillage or air embolism USFDA / European CE/ Approved / Equivalent Indian Standard Approved	14.5 fr. X 23 cm
612	Tunnelled Cuffed Double Lumen Chronic Dialysis Catheter: straight(adult)	USFDA / European CE/ Approved / Equivalent Indian Standard Approved Sterile pack Biocompatible Sterile Kink resistant Material should be biocompatible, Complete sterile pack including complete kit. Chronic HD catheter with symmetric tip design improve patency and flow as high as 500 ml/min design, size 14.5 Fr size, less than 1% recirculation rate Tapered Cuff, and 360 degree side holes with air guard.Should have preloaded hollow stylet in order to have ease of insertion in straight catheters. Catheter type stratight	Size: 33 cm
613	Soft Cannula	Cannula with Universal standard connector, Lightweight, flexible nasal sof-tech special soft cannula with anatomical design must be with channelled star inner lumen, crush & kink resistant tubing. Must be USFDA / BIS / AUS/ EU CE certified with ICMED 13485 or ISO13485 Certification.	
614	Excerciser	Volumetric exerciser with inbuilt ergonomic handle, MP holder, 2500 ml inspiratory capacity, range 250 to 2500 ml flow rate, Calibrations visible on both sides, bright green Flow indicator piston with inspiring smiley face imagery with Green Marker to set inspiratory goal. Polystyrene & Polypropylene (PP) construction, Particulate filter screen in device housing, Ergonomic, swivel (PP) mouthpiece with Expandable collapsible (PP) tube. Must be USFDA / BIS / CE Certified with ICMED 13485 or ISO13485 Certificate.	
615	Excerciser	Volumetric exerciser with inbuilt ergonomic handle, MP holder, 4000 ml inspiratory capacity, range 250 tp 4000 ml flow rate, Calibrations visible on both sides, bright green flow indicator piston with inspiring smiley face imagery with Green Marker to set inspiratory goal. Polystyrene & Polypropylene (PP) construction, Particulate filter screen in device housing, Ergonomic, swivel (PP) mouthpiece with Expandable collapsible (PP) tube. Must be USFDA / BIS / CE Certified with ICMED 13485 or ISO 13485 Certification.	
616	Marker	SKIN MARKER TIP WITH RULER, VIOLET INK Sterile Medical Grade, non-irritating, violet ink skin marker with ruler, USFDA / CE / BIS, ISO 13485	



# ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

\*\*\*\*\*

S. No.	Item Name	Specification	Size
617	Sodium Acrylate Cover [Pre Intra & Post OP without machine]	<ol style="list-style-type: none"><li>1. It can be used before, during and after surgery to help prevent peri-hypothermia in patients.</li><li>2. It should be self-warming and work without additional equipment / electric supply for setting up and during use.</li><li>3. The blanket should begin warming immediately once removed the packaging and should reach operational temperature (average 40 degree C) within 30 minutes, which is maintained for several hrs.</li><li>4. It should be made of Several layers of SMS (polypropylene). The layers and pockets should be sewn together with polyester / cotton threads.</li><li>5. It should have 12 warming pads that are made up of air-permeable bags.</li><li>6. Should be 4 digit European CE notified body / BIS.</li></ol>	
618	Mask with Anti Spill Jar	Aerosol Mask Kit Pediatric / Adult, with high-performing chamber designed to deliver a fine, dense mist at a particle size of 2.1µm at all angles up to 90° to maximize placement allowing flexible medication delivery, Easy-seal, threaded, cap, 6cc capacity, Jet to stay in place. Anti-spill jar to prevent leakage with recommended flow rate of 8 LPM to be supplied with non-kinking star lumen tubing with standard / universal / Clear Connect Secure Tubing Connector for Nebulization rate of > 20 mL / min. US FDA registered & if applicable 510K. Must be USFDA / BIS / CE Certified with ICMED 13485 or ISO 13485 Certification.	
619	T Piece with Mask with Anti Spill Jar	Aerosolization Mouthpiece & Tee Connector Kit, with high-performing chamber designed to deliver a fine, dense mist at a particle size of 2.1µm at all angles up to 90° to maximize placement allowing flexible medication delivery, Easy-seal, threaded, cap, 6 cc capacity, Jet to stay in place. Anti-spill jar to prevent leakage with recommended flow rate of 8 LPM to be supplied with non-kinking star lumen tubing with standard / universal / Clear connect Secure Tubing Connector for Nebulization rate of >20 mL / min. US FDA registered & if applicable 510K. Must be USFDA / BIS / CE Certified with ICMED 13485 or ISO 13485 Certification.	

**Note:**

1. Above mentioned modifications will not reflect in BoQ, therefore, **prices for newly added items can be filled after inserting New Sheet in BoQ.**

**Please refer Corrigendum > Items added in Existing List >**

**Following items are deleted from existing list:**

S. No.	Item Name	Specification	Size
585	Surgical Mask	Three Ply Mask (without elastic band)– First Layer – Meltblon 175 mm, thickness 25 GSM, Medium Layer blue thickness 25 GSM, 175 mm Last Layer – White Thickness 25 GSM, 190 mm. The tie patti 32 GSM &	Mask size 7" x 4"

**Executive Director**