



अखिल भारतीय आयुर्विज्ञान संस्थान, जोधपुर
ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

PROC-2/RC/18/2024-AIIMS.JDH

Dated: 02nd May, 2025

CORRIGENDUM

For

RATE CONTRACT

FOR SUPPLY OF

Implantable Hearing Devices

All India Institute of Medical Sciences, Jodhpur

NIT No.	:	PROC-2/RC/18/2024-AIIMS.JDH
NIT Issue Date	:	13 th March, 2025
Revised Last Date of Submission	:	13 th May, 2025 (3:00 PM)
Date of Opening	:	14 th May, 2025

Please refer tender document:

			<u>For</u>	<u>Read As</u>
<u>General Terms & Conditions</u>				
1	Page No. 13	<u>37. LIQUIDATED DAMAGES</u>	Supply of material will have to be completed within the delivery period or period mentioned in the purchased order. The liquidated damages charges @ 0.5% per week shall be imposed if supply made after expiry of delivery period subject to maximum 10% of the total value of relevant goods. Quantum of liquidated damages assessed and levied by the purchaser shall be final and not challengeable by the supplier	<p>(i) Supply of material will have to be completed within the delivery period or period mentioned in the purchased order. The liquidated damages charges @ 0.5% per week shall be imposed if supply made after expiry of delivery period subject to maximum 10% of the total value of relevant goods. Quantum of liquidated damages assessed and levied by the purchaser shall be final and not challengeable by the supplier.</p> <p>(ii) Non supply of items shall attract a penalty of 10% of the value of non-supplied goods/material.</p> <p>(iii) The penalty can be waived off by the competent authority in emergent/extraordinary situation.</p>
2	Page No. 13	<u>39. REPLACEMENT OF EXPIRED ITEM</u>	<p>a) Expired items will be exchanged within Two Days by the supplier with fresh lot as informed by the Department / Central Store. No communication in this regard will be entertained regarding timely information / prior to expiry of items.</p> <p>b) If the supplier fails to exchange the expired goods / items within the timeframe, it will be termed as breach of contract conditions, AIIMS Jodhpur reserves the right to initiate any suitable action against the supplier in this regards.</p> <p>(c) The supplier shall arrange to effect free replacement of any quantity which may deteriorate in potency, strength approaching expiry or expired etc. before the date of expiry marked on the labels.</p>	<p>(a) Expired/ Near Expiry items will be exchanged within Two weeks by the supplier with fresh lot as informed by the Department / Central Store. No communication in this regard will be entertained regarding timely information / prior to expiry of items.</p> <p>(b) If the supplier fails to exchange the expired goods / items within the timeframe, it will be termed as breach of contract conditions, AIIMS Jodhpur reserves the right to initiate any suitable action against the supplier in this regard.</p> <p>(c) The supplier shall arrange to effect free replacement of any quantity which may deteriorate in potency, strength approaching expiry or expired etc. before the date of expiry marked on the labels.</p>

General Terms for all the above eight categories:

3	Page 35	1	In addition to the quoted models against the technical specifications given above for the purpose of open tendering and selection of L1 bid, please provide the list and proposed rates of all the available models and their combinations for the purpose of procurement through AIIMS Jodhpur, in case the Institute needs.	In addition to the quoted models against the technical specifications given above for the purpose of open tendering and selection of L1 bid, please provide the list and proposed rates of all available models, their spares, accessories and consumables, Audio Processor Upgrade Kits, and Passive Middle Ear Implants for the purpose of procurement through AIIMS Jodhpur, in case the Institute needs. The list may be given as an additional separate sheet in the Price Bid BOQ Excel File.
4		4	The quoted models and company to be in existence and held for minimum for 5 years in India	<u>Deleted</u>

ANNEXURE-XIV

5	Page 30	S. No. 1	Item Code: 1.001	Basic Cochlear Implant with BTE Speech Processor	i) Features of Implants	d) Should be MRI Compatible at 1.5 Tesla without removal of magnet should be compatible as per USA Standard	Should be MRI compatible at 1.5 Tesla without removal of magnet, should be compatible as per USA standard. If any kit is required to be used during MRI at 1.5 Tesla without removal of magnet, the kit should be provided free of cost to the patient whenever it is required, for the period of warranty of the internal implant. 10 kits should always be kept with AIIMS Jodhpur and expired ones should be replaced immediately.
6					ii) Features of Speech Processor	d) should have wide input dynamic range (minimum 75dB), dual microphone with automatic adaptive directionality feature for better understanding of speech in noise	Should have features for better understanding of speech in noise
7	Page 31	S. No. 2	Item Code: 1.002	Basic Cochlear Implant with single unit speech	(i) Features of implant	d) Should be MRI Compatible at 1.5 Tesla without removal of magnet should be compatible as per USA Standard	Should be MRI compatible at 1.5 Tesla without removal of magnet, should be compatible as per USA standard. If any kit is required to be used during MRI at 1.5 Tesla without removal of magnet, the kit should be provided free of cost to the patient whenever it is required, for

				processor			the period of warranty of the internal implant. 10 kits should always be kept with AIIMS Jodhpur and expired ones should be replaced immediately.
8					(ii) Features of Speech Processor	e) should have wide input dynamic range (minimum 75dB) with automatic feature for better understanding of speech in noise	Should have features for better understanding of speech in noise
9	Page 32	S. No. 3	Item Code: 1.003	Advanced Cochlear Implant with BTE speech processor	(ii) Features of Implants	d) Should be MRI Compatible at 3.0 Tesla without removal of magnet, should be compatible as per USA Standard	Should be MRI compatible at 1.5 Tesla without removal of magnet, should be compatible as per USA standard. If any kit is required to be used during MRI at 3 Tesla without removal of magnet, the kit should be provided free of cost to the patient whenever it is required, for the period of warranty of the internal implant. 10 kits should always be kept with AIIMS Jodhpur and expired ones should be replaced immediately
10					(iii) Features of Speech Processor	b) should be moisture/water resistant, should have minimum IP54 rating	Should be moisture/water resistant
11						e) should have wide input dynamic range (minimum 75dB) with automatic features for better understanding of speech in noise. Dual microphones with automatic adaptive directionality required	Should have advanced features for better understanding of speech in noise
12	Page 33	S.No. 4	Item Code: 1.004	Advanced Cochlear Implant with Single Unit Processor	ii) Features of Implants	d) Should be MRI Compatible at 3.0 Tesla without removal of magnet should be compatible as per USA Standard	Should be MRI compatible at 1.5 Tesla without removal of magnet, should be compatible as per USA standard. If any kit is required to be used during MRI at 3 Tesla without removal of magnet, the kit should be provided free of cost to the patient whenever it is required, for the period of warranty of the internal implant. 10 kits should always be kept with AIIMS Jodhpur and expired

						ones should be replaced immediately.
13					(iii) Features of Speech Processor	b) should be moisture/water resistant, should have minimum IP54 rating
14						e) should have wide input dynamic range (minimum 75dB) with automatic features for better understanding of speech in noise. Dual microphones with automatic adaptive directionality required

Executive Director