



Date: -15th November, 2017

Corrigendum
For
Tender for
Semi-Automated ELISA Reader with washer for the
Department of Microbiology

NIT Issue Date	: 09 th October, 2017
NIT No.	: Admn/Tender/212/2017-AIIMS.JDH
Pre-Bid Meeting	: 24 th October, 2017 at 03:30 PM
Last Date of Submission	: 27 th November, 2017 at 03:00 PM
Bid opening	: 28 th November, 2017 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Point No. 01:

For

Should have 8 channel single beam absorption photometer with auxiliary beam path & discontinuous wavelength feed.

Read

Should have 8 channel optical system with single, dual and multichromatic (up to 3) wave length selection.

2. Page No. 10, Point No. 03:

For

Should be able to evaluate colorimetric endpoint specification and kinetic tests.

Read

Should be able to calculate: Nonlinear standard, cut off, index formula, column subtraction, linear regression, point to point, cube spline, linear-log.

3. Page No. 10, Point No. 04:

For

Should be able to control the temperature of the ELISA plate to 30°C-37°C.

Read

Should be able to control the temperature of the ELISA plate to 30°C-37°C with variable shaking facility.