



Date: - 21st June, 2018

Corrigendum
For
Neuro OT Table for the Department of Neurosurgery

NIT Issue Date	: 07 th May, 2018
NIT No.	: Admn/Tender/228-2/2017-AIIMS.JDH
Pre-Bid Meeting	: 17 th May, 2018 at 03:15 PM
Earlier Last Date of Submission	: 28 th June, 2018 at 03:00 PM
Extended Last Date of Submission	: 04 th July, 2018 at 03:00 PM
Bid opening	: 05 th July, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 11, 1st Sub point of “Table Top” :

For

Dimensions: Minimum 450 mm* 1850 mm

Read

Dimensions: Minimum 500 mm* 2120 mm

2. Page No. 11, 2nd Sub point of “Table Top” :

For

Adjustable Height: Minimum 400 – 550 mm, Maximum above 1000mm

Read

Adjustable Height: Minimum 520 mm, Maximum above or equal to 1000 mm

3. Page No. 11, 3rd Sub point of “Table Top” :

For

Tilt

- Trendelenburg/Reverse Trendelenburg 25-45 degree
- Right and Left Lateral Tilt 20 degrees
- Back Section -30 to +90
- Head section -15 to +90 degrees
- Leg section -45 to -90

Read

Tilt

- Trendelenburg/Reverse Trendelenburg 30-40 degree
- Right and Left Lateral Tilt 35 degrees
- Back Section 40 to 90
- Head section 60 to 90 degrees
- Leg section -90 degrees

4. Page No. 11, 4th Sub point of “Table Top” :

For

Table top sliding facility min 53 cms of top slide and return – to – level

Read

Table top sliding facility min 65 cms towards leg and head section.

5. Page No. 11, 6th Sub point of “Table Top” :

For

Construction material – transparent radiolucent acrylic

Read

Construction material – Stainless Steel

6. Page No. 11, 23rd Sub point of “Accessories” :

For

Brain spatula

- 6 mm x 2
- 4 mm x 2
- 2 mm x 2

Read

Brain spatula

- 10 mm x 2
- 8 mm x 2
- 6 mm x 2
- 4 mm x 2
- 2 mm x 2

7. Page No. 12, 28th Sub point of “Accessories”:

For

Wilson Frame Radiolucent

Read

Deleted

8. Page No. 12, 29th Sub point of “Accessories”:

For

Prone position helmet

Read

Prone position helmet with disposable face mask 40 in number.

9. Page No. 12, Addition of Point under Heading : “Conditions”

Sugita Head Frame and its related accessories should be CE (Europe) and USA FDA Approved and by the same manufacturer as that of the table.