



Date: - 29th June, 2018

Corrigendum
For
Tender for
Laparoscopic Instruments – General for the
Department of Surgical Gastroenterology

NIT Issue Date	: 09 th April, 2018
NIT No.	: Admn/Tender/27/2018-AIIMS.JDH
Pre-Bid Meeting	: 25 th April, 2018 at 03:00 PM
Earlier Last Date of Submission	: 25 th June, 2018 at 03:00 PM
Extended Last Date of Submission	: 09 th July, 2018 at 03:00 PM
Bid opening	: 10 th July, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 11, 3rd Sub point of Scissors:

For

Autoclavable Laparoscopic Scissors Spoon shaped, Double action Jaw Serrated Jaws 5mm

Read

Autoclavable Laparoscopic curved Scissors, Double action Jaw Serrated Jaws 5mm

2. Page No. 11, 1st subpoint of clip applicator:

For

Reusable Endo Clip Applier With Rotating Shaft For Medium Clips (Lt 200).

Read

Reusable Endo Clip Applier With Rotating Shaft For Medium Clips.

3. Page No. 11, 1st subpoint of clip applicator:

For

Reusable Endo Clip Applier With Rotating Shaft For Medium Large Clips (Lt 300).

Read

Reusable Endo Clip Applier With Rotating Shaft For Medium Large Clips .

4. Page No. 11, 3rd subpoint of clip applicator

For

Reusable Endo Clip Applier with Rotating Shaft For Large Clips (Lt 400)

Read

Reusable Endo Clip Applier with Rotating Shaft for Extra Large Clips.

5. Page No.11, 4th sub point of clip applicator

For

Reusable Hemostatic Clip Applier For Laparoscopic Surgery For Application Of Hemostatic Locking Non Absorbable Clips Of Sizes Medium Large, Large And Extra Large, 5 Mm Diameter 32 Cm Curved

Read

Reusable Hemostatic Clip Applier For Laparoscopic Surgery For Application Of Hemostatic Locking Non Absorbable Clips Of Sizes Medium Large, Large And Extra Large, 5 Mm Diameter 32 cm

6. Page No. 11, 6th sub point of clip applicator

For

Bipolar High Frequency Cord Compatible With Bipolar Electro Surgical Generator And Bipolar Laparoscopic Hand Instruments

Read

No electrosurgical generator required only compatible Bipolar cord and bipolar laparoscopic hand instruments

7. Page No. 12, 9th sub point of Trocars

For

Disposable Opaque Body Bladeless Trocar With Converter Less Insertion Of Instruments With Different Diameters, Having Insuffulation Port And Ribbed Cannula, With Obturator Having Safety Shield. 5 Mm, 75/100 Mm Length

Read

Disposable Body Bladeless Trocar With Converter Less Insertion Of Instruments With Different Diameters, Having Insuffulation Port And Ribbed Cannula, With Obturator Having Safety Shield. 5 Mm, 75/100 Mm Length

8. Page No. 12, 11th sub point of Trocars

For

Disposable Opaque Body Bladeless Trocar With Converter Less Insertion Of Instruments With Different Diameters, Having Insuffulation Port And Ribbed Cannula, With Obturator Having Safety Shield. 5-12 Mm, 75/100 Mm Length

Read

Disposable Body Bladeless Trocar With Converter Less Insertion Of Instruments With Different Diameters, Having Insuffulation Port And Ribbed Cannula, With Obturator Having Safety Shield. 5-12 Mm, 75/100 Mm Length

9. Page No. 12, Addition of paragraphs

Haemostatic locking non absorbable clips medium large, large, extra large

10. Page No. 12, Addition of paragraphs

Laparoscopic Reusable haemostatic lock clip remover. Quantity 1

11. Page No. 12, Addition of paragraphs

The all instruments mentioned should be quoted and the 80% must be from the same manufacturer and 20% may from 3rd party but the responsibility of the warranty will be of the bidder and all instruments should be as per quality standard asked in the tender.

12. Page No. 12, Addition of paragraphs

Instruments should be European CE (with four digit notified body no.) or US FDA for the quoted instruments.