



Date: - 03rd August 2016

Corrigendum

for

Mayo's Trolley

NIT No.	: Admn/Tender/33/2016-AIIMS.JDH
NIT Issue Date	: 23 rd June, 2016
Pre Bid Meeting	: 04 th July, 2016 at 05:10 PM
Last Date of Submission	: 22 nd July, 2016 at 03:00 PM
Extended Last Date of Submission	: 10 th August, 2016 at 03:00 PM
Revised Last Date of Submission	: 23 rd August, 2016 at 03:00 PM

1. The following revised and additional specification will be added:-

1. Page No. 10, Chapter-II Specification and Allied Technical Specification, Parameters and Technical Specifications for Executing the Work:, Point No. 01:

For

All dimensions shall be read in metric unit only. Tolerance in overall dimensions upto $\pm 5\%$. No other deviations from specifications shall be allowed.

Read

All dimensions shall be read in metric unit only. Tolerance in overall dimensions upto -2% to $+5\%$. No other deviations from specifications shall be allowed.

2. Page No. 10, Chapter-II Specification and Allied Technical Specification, Parameters and Technical Specifications for Executing the Work:, General Eligible Criteria, Point No. 02:

For

Manufacturer should have valid European CE, ISO 9001:2008, ISO 14001:2004, and OHSAS 18001:2007 certificate. Dealers participating should enclose certificate from their parent manufacturer company.

Read

Manufacturer should have valid CE, ISO 9001:2008, ISO 14001:2004, and OHSAS 18001:2007 certificate. Dealers participating should enclose certificate from their parent manufacturer company.

3. Page No. 14, Chapter-V, TECHNICAL SPECIFICATION FOR MAYO'S TROLLEY, Bullet No. 01:

For

Adjustable height of SS tray from 845 mm to 1300 mm

Read

Height Adjustable SS tray with minimum height from 730mm to 845mm and maximum height from 1280mm to 1320mm. The general variation of -2% to $+5\%$ is not applicable on this.

**Administrative Officer
AIIMS, Jodhpur**