



Date: - 10<sup>th</sup> September, 2018

Corrigendum  
For  
High End Cardiac Monitors for the Department of  
Cardiology

NIT Issue Date	: 14 <sup>th</sup> August, 2018
NIT No.	: Admn/Tender/37-2/2018-AIIMS.JDH
Pre-Bid Meeting	: 24 <sup>th</sup> August, 2018 at 03:00 PM
Earlier Last Date of Submission	: 20 <sup>th</sup> September, 2018 at 03:00 PM
Extended Last Date of Submission	: 26 <sup>th</sup> September, 2018 at 03:00 PM
Bid opening	: 27 <sup>th</sup> September, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 11, Point No. 2 :

**For**

Monitor should have 19” independent flat panel display. Should be capable of 8 traces display

**Read**

Monitor should have 19” Independent flat panel display & should be IT enable & should host all third party application like HIS, PACS, LIS, CIS on the same screen simultaneously.

2. Page No. 11, Point No. 4 :

**For**

Monitor must be capable of simultaneously monitoring the following parameters which should be present as standard: ECG, NIBP, SpO2, invasive pressures (4), temperatures (2), Capnography and cardiac output monitoring.

**Read**

Monitor must be capable of simultaneously monitoring the following parameters which should be present as standard: ECG, NIBP, SpO2 (Masimo rainbow technology) It should be upgradeable to PVI, PI, SPOC, SPCO, invasive pressures (4), temperatures (2), Capnography (ETCO2 sample collection must be 50 ml/min) and cardiac output monitoring.

3. Page No. 11, Point No. 5 :

**For**

Should be compatible with Cardiac output, EEG, and BIS.

**Read**

Should be compatible with EEG, and BIS.

**4. Page No. 11, Point No. 6 :**

**For**

ECG should have capability for 3, 5 and / or 10 lead monitoring and should have built in arrhythmia monitoring on all leads.

**Read**

ECG should have capability for 10 lead monitoring and should have built in arrhythmia monitoring on all leads.

**5. Page No. 11, Point No. 10 :**

**For**

Respiration should be available with Cardio Vascular Artifact filter.

**Read**

Respiration should be available with latest filter technology for artifact removal.

**6. Page No. 12, Point No. 33 :**

**For**

Complete monitoring system should have US FDA/ European CE certifications.

**Read**

Complete monitoring system should have US FDA and European CE certifications.

**7. Page No. 12, In Central Monitoring Station for multi para monitor, Point No. 1 :**

**For**

System should have minimum 16 beds capability.

**Read**

System should have minimum 16 beds capability and two separate central monitoring stations are needed

**8. Page No. 12, In Central Monitoring Station for multi para monitor, Point No. 2 :**

**For**

System should have minimum 32 beds capability and two separate stations are needed.

**Read**

Central station should have 21" color display with 4 days trends.

**9. Page No. 12, In Central Monitoring Station for multi para monitor, Point No. 5 :**

**For**

Should have separate computer keyboard and network laser printer.

**Read**

Should have separate latest generation computer, wireless keyboard and network laser printer. The provided computer system including accessories and printers should be of latest technology available in the market.

**10. Page No. 12, In Central Monitoring Station for multi para monitor, Last Para:**

**For**

Should have capability for HL7 interface. Each bedside monitor should have compatible transport module or transport monitor. All systems should have US-FDA or European CE certifications.

**Read**

Should have capability for HL7 interface. Each bedside monitor should have compatible transport module or transport monitor. All systems should have US-FDA and European CE certifications.