



Date: - 13<sup>th</sup> February 2017

## Corrigendum

for

### Automated ELISA readers with washer for the Department of Microbiology

NIT Issue Date	:	03 <sup>rd</sup> October 2016.
NIT No.	:	Admn/Tender/61/2016-AIIMS.JDH
Pre Bid Meeting held on	:	17 <sup>th</sup> October, 2016 at 11:30 AM
Last Date of Submission	:	22 <sup>nd</sup> November, 2016 at 03:00 PM
Revised Last Date of Submission	:	09 <sup>th</sup> February, 2017 at 03:00 PM
Extended Last Date of Submission	:	01 <sup>st</sup> March, 2017 at 03:00 PM

#### **The following revised and additional specification will be added:-**

**1. Page No. 10, In technical specification, Point No. 09:**

**For**

Machine should be supplied with 6 standard filters.

**Read**

Machine should be supplied with 6 filters of user's choice.

**2. Page No. 10, In technical specification, Point No. 21:**

**For**

User's list should be attached with satisfactory report for the last three years from three blood bank users with contact details.

**Read**

User's list should be attached with satisfactory report for the last three years from the users from any three Institutes with contact details.

**3. Page No. 10, In technical specification, Point No. 23:**

**For**

Centrifuge should be FDA approved or European CE

**Read**

Elisa Reader should be FDA approved or European CE

**4. Page No. 10, In technical specification, In Elisa Plate Washer, Point No. 09:**

**For**

Should have in-built vacuum and dispensing pumps to ensure accurate and quite washing.

**Read**

Should have in-built vacuum/ Non Vacuum/ Purostatic and dispensing pumps to ensure accurate and quite washing.

**5. Page No. 11, In technical specification, In Elisa Plate Washer, Point No. 15:**

**For**

Firm will have to supply compatible UPS with minimum half hr backup along with the equipment free of cost.

**Read**

Deleted.

**6. Page No. 10, In technical specification, In Elisa Reader, Bullet No. 03:**

**For**

Optical systems: LED lamp/ UV Xenon flash lamp

**Read**

Optical systems: LED lamp/ Xenon flash lamp

**Administrative Officer  
AIIMS, Jodhpur**