



Date: - 18th February, 2020

Corrigendum
for
High End Stretcher Trolley for Modular Operation
Theater and ICU

NIT Issue Date	: 09 th January, 2019
NIT No.	: Admn/Tender/62/2019-AIIMS.JDH
Pre-Bid Meeting	: 21 st January, 2020 at 03:15 PM
Earlier Last Date of Submission	: 17 th February 2020 at 03:00 PM
Extended Last Date of Submission	: 28 th February 2020 at 03:00 PM
Bid opening	: 29 th February, 2020 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Point No. 02

For: -

X-Ray Stretcher should have dual-deck design allows cassette to be adjusted to the patient.

Read: -

X-Ray stretcher should have design to allow cassette to be adjusted to the patient.

2. Page No. 10, Point No. 05

For

It should have thermoformed ABS base with extra capacity storage tray

Read

It should have high pressure laminate/thermoformed ABS base tray with extra storage capacity.

3. Page No. 10, Point No. 06

For

It should have 700 lb. weight capacity to transfer bariatric or obese patients.

Read

It should have 550 lb. (>200KG) weight capacity to transfer bariatric or obese patients.

4. Page No. 10, Point No. 07

For

It should have less than 21-inch low height for easy egress of patient from stretcher.

Read

It should have less than 24-inch low height for easy egress of patient from stretcher.

5. Page No. 10, Point No. 10

For

It should have integrated transfer board for lateral patient transfer to OT table or beds with out Patient lift or glide sheets.

Read

It should be supplied with Integrated/separate transfer board for lateral patient transfer to OT table of Bed.

6. Page No. 10, Point No. 17**For**

It should have 8-inch Omni Surgace™ Casters with integrated wheel covers.

Read

It should have 8-inch casters with integrated wheel covers.

7. Page No. 10, Point No. 24**For**

Overall Length 85" (216 cm)

Read

Range Overall length 85"-90" inch.

8. Page No. 10, Point No. 25**For**

Overall width side rails up 38" (96.5 cm)

Read

Overall width side rails up 30"-38" inch.

9. Page No. 10, Point No. 26**For**

Overall width side rails down 30.5" (77.5 cm)

Read

Overall width side rails down 28"-32" inch.

10. Page No. 10, Point No. 28**For**

Weight capacity 700lb. (318kg)

Read

Weight capacity more than 200kg

11. Page No. 10, Point No. 29**For**

Height range high 37.25" (94.6 cm)

Read

Height range high 35"-40" inch

12. Page No. 10, Point No. 30**For**

Height range low 24" (61 cm)

Read

Height range low 24" inch

13. Page No. 10, Point No. 31**For**

Litter Positioning Backrest 0⁰-90⁰

Read

Change to Litter Positioning backrest 0⁰-85 to 90⁰

14. Page No. 10, Point No. 33**For**

Trend/Reverse Trend +- 17⁰

Read

Trend/Reverse Trend +- 15⁰ or better