



Date: - 13th April, 2020

Corrigendum
for
Radiant Warmer for the Department of Pediatrics

NIT Issue Date	: 07 th February, 2020
NIT No.	: Admn/Tender/71/2019-AIIMS.JDH
Pre-Bid Meeting	: 18 th February, 2020 at 04:45 PM
Last Date of Submission	: 15 th April, 2020 at 03:00 PM
Last Date of Submission	: 29 th April, 2020 at 03:00 PM
Bid opening	: 30 th April, 2020 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page No. 11, Point –“ 9 (l)”**
For
Should have LCD graphical display with the facility of trending temperature.
Read
Should have LCD graphical display with the facility of trending temperature (Optional)
- 2. Page No. 12, point no. “31 (a)”**
For:
Reusable thermistor temperature probes (For measuring core temperature): 05 (Five) per radiant warmer.
Read:
Reusable thermistor temperature probes (For measuring core temperature): 05 (Five) per radiant warmer
Or
200 disposable probes per radiant warmer
- 3. Page No. 13, point no. 31 (b):**
For:
Reusable thermistor temperature probes (For measuring periphery temperature): 05 (Five) per radiant warmer
Read:
Reusable thermistor temperature probes (For measuring periphery temperature): 05 (Five) per radiant warmer
Or
200 disposable probes per radiant warmer