



Date: - 01st February, 2017

Corrigendum
for
tender
for
ACT (Activated Clothing Time) Machine for the
Department of Nephrology

NIT Issue Date	:	18 th November 2016
NIT No.	:	Admn/Tender/84/2016-AIIMS.JDH
Pre Bid Meeting held on	:	30 th November, 2016 at 03:00 PM
Earlier Date of Submission	:	07 th February, 2017 at 03:00 PM
Extended Last Date of Submission	:	14 th February, 2017 at 03:00 PM
Bid Opening	:	15 th February, 2017 at 11:30 AM

The following revised and additional specification will be added:-

1. Page No. 11, In technical specification, Toggle Button No. 12:

For

Guarantee / Warrantee Period: The Tenderers must quote for **05 years** comprehensive warranty (Including all Spares, Accessories and Labour) from the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. Also the bidders are requested to submit their quote (Rates) for subsequent **05 years** Comprehensive Maintenance Contract (CMC) (Including All Spares, Accessories and Labour). Failure to comply this condition will entail the rejection of the bids. The price comparison shall be taking into account on basic price and post warranty CMC.

Read

Guarantee / Warrantee Period: The Tenderers must quote for **02 years** comprehensive warranty (Including all Spares, Accessories and Labour) from the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. Also the bidders are requested to submit their quote (Rates) for subsequent **03 years** Comprehensive Maintenance Contract (CMC) (Including All Spares, Accessories and Labour). Failure to comply this condition will entail the rejection of the bids. The price comparison shall be taking into account on basic price and post warranty CMC.

2. Page No. 11, In technical specification, Toggle Button No. 16:

For

Dual well testing method.

Read

Deleted

- 3. Page No. 11, In technical specification, Toggle Button No. 17:**
For
Accepts actalyte ACT tubes with celite, glass bead activator, MAX ACT tubes with blended activator; all international technidyne Hemochron tubes.
Read
Deleted
- 4. Page No. 11, In technical specification, Toggle Button No. 21:**
For
It should have a battery backup of 2 hrs
Read
It should have a inbuilt battery/ UPS power backup of 2 hrs.
- 5. Page No. 11, In technical specification, After Toggle Button No. 23:**
Added Point No. 24:
High performance proven technology clot detection system should be there
- 6. Page No. 11, In technical specification, After Toggle Button No. 24:**
Added Point No. 25:
Should have quality control manager and quality control menu which should allow daily testing and verification of ACT plus readings.

Administrative Officer
AIIMS, Jodhpur