



Date: - 28th January 2019

Corrigendum
For
Surgical Micro Drill & Saw System for the
Department of Dentistry

| | |
|----------------------------------|--|
| NIT Issue Date | : 22 nd October, 2018 |
| NIT No. | : Admn/Tender/88/2018-AIIMS.JDH |
| Pre-Bid Meeting | : 30 th October, 2018 at 03:00 PM |
| Eelier Last Date of Submission | : 30 th January, 2019 at 03:00 PM |
| Extended Last Date of Submission | : 14 th February, 2019 at 03:00 PM |
| Bid opening | : 15 th February, 2019 at 03:15 P.M |

The following revised and additional specification will be added:-

- Page 10, Point 1 “Console” sub-point K :**
For
On Screen Help
Read
Deleted
- Page 10, Point 3 “Drill Handpiece” sub-point A :**
For
Maximum Speed > 75000 RPM
Read
Maximum Speed 50000 to 75000 RPM.
- Page 10, Point 5 “Burs for the Attachments” sub-point A :**
For
 - Assorted Tungsten Carbide Cutting & Diamond from 1 mm to 7 mm
 - 10 Burs of each size**Read**
 - Assorted Tungsten Carbide Cutting & Diamond from 2 mm to 7 mm
 - 3 to 4 Burs of each size
- Page 10, Point 6 “Micro saws”:**
For
 - Sagittal Saw
 - Maximum speed of 23000 cpm
 - Oscillating Saw
 - Maximum speed of 22000 cpm
 - Reciprocation Saw
 - Maximum Speed of 18000 cpm

Read

- a. Sagittal Saw
 1. Maximum speed of 20000 to 25000 cpm
- b. Oscillating Saw
 1. Maximum speed of 20000 to 25000 cpm
- c. Reciprocation Saw
 1. Maximum Speed of 14000 cpm to 18000 cpm

5. Page 11, Point 6 “Wire and Pin Driver”, sub point A:

For

- a. Maximum speed of 1500 rpm

Read

- a. Maximum speed 1200 to 1500 rpm

6. Page 11, Point 12, “Connecting Cord”:

For

- a. 10ft long, 3/8" diameter flexible electrical connecting cord
- b. Dot-to-Dot type push-pull connectors at both ends.
- c. Autoclavable

Read

- a. 10ft long, 3/8" diameter flexible electrical connecting cord
- b. Dot-to-Dot type push-pull connectors at both ends.
- c. Autoclavable OR The connecting Cable/Cord should be integrated and autoclavable.

7. 14th Point to be added after 13th :

Hand piece lubrication and cleaning solution 06 Nos.

8. 15th Point to be added after 14th

Irrigation spray nozzle should be supplied with all hand piece attachments.

9. 16th Point to be added after 15th :

Hand piece lubrication and cleaning solution should be USFDA or CE approved