



Date: - 12th February 2016

Corrigendum
for
Equipments required for Department of
Anaesthesia

NIT Issue Date	:	15 th January 2016.
NIT No.	:	Admn/Tender/Anaesthesia/2015-AIIMS.JDH
Pre Bid Meeting held on	:	29 th January,2016 at 03:00 AM
Last Date of Submission	:	16 th February, 2016 at 03:00 PM
Revised Last Date of Submission	:	29 th February, 2016 at 03:00 PM

1. The following revised and additional specification will be added:-

- 1. Page No. 11, Item No. 2, Portable Colour Doppler Ultrasound Machine with 2D Echo, Point No. 12,**
For
Flat LCD/ TFT monitor of at least 15 inches with flicker free image.
Read
Flat LCD/ TFT monitor of at least 10 inches with flicker free image.
- 2. Page No.11, Item No. 2, Portable Colour Doppler Ultrasound Machine with 2D Echo, Point No. 13:**
For
Alphanumeric soft keys keyboard with easy access scans controls.
Read
Alphanumeric soft keys keyboard with easy access scans controls and facility to sanitize the system including key board to avoid cross-contamination.
- 3. Page No. 12, Item No.4, Patient Warming System, Point No. 2**
For
Should have single hose for all type/size of blankets
Read
Should have corrugated hose pipe for all type/size of blankets

4. Page No. 12, Item No.4, Patient Warming System, Point No. 5

For

Should have digital hour meter.

Read

Should have LCD display of temperature setting with (+/- 1% accuracy) of displayed temperature.

**Administrative Officer
AIIMS, Jodhpur**