



Date: - 12<sup>th</sup> February 2016

**Corrigendum**  
**for**  
**Equipments required for Department of**  
**Blood Bank**

NIT Issue Date	:	15 <sup>th</sup> January 2016.
NIT No.	:	Admn/Tender/Blood Bank/2015-AIIMS.JDH
Pre Bid Meeting held on	:	29 <sup>th</sup> January 2016 at 11:00 AM
Last Date of Submission	:	16 <sup>th</sup> February 2016 at 03:00 PM
Revised Last of Submission	:	26 <sup>th</sup> February 2016 at 03:00 PM

**1. The following revised and additional specification will be added:-**

**1. Page 10, Item 01, Semi automated coagulometer, Point No. 4:**

**For:**

At least 2-4 positions for reagents of which one should be with magnetic stirrer”

**Read:**

At least 2-4 positions for reagents .

**2. Page 10, Item 01, Semi automated coagulometer, Point No. 5:**

**For**

At least 2 pipette wells

**Read**

Deleted

**3. Page 10, Item 01, Semi automated coagulometer, Point No. 10:**

**For:**

Should give results in Seconds and in various calculated units (% , INR, Ratio, Gm/L, mg/dL, IU/ml).

**Read:**

Should give results in Seconds and in various calculated units (% , INR, Ratio, Gm/L, mg/dL).

**4. Page 10, Item 01, Semi automated coagulometer, Point No. 14:**

**For:**

One complete set of accessories should be provided as a start up which should include:

**Read:**

One complete set of accessories should be provided which should include:

**5. Page 10, Item 01, Semi automated coagulometer, Point No. 14, Sub Point No. a:**

**For:**

Measurement Cuvettes: 1000 Pieces.

**Read**

Measurement Cuvettes with corresponding magnetic bead/stirrer: 1000 Pieces.

**6. Page 10, Item 01, Semi automated coagulometer, Point No. 14, Sub Point No. b:**

**For:**

Stage Autopipette: 1 Piece (25/50/100/200µl)

**Read**

Compatible calibrated micropipette: 1 Piece (25/50/100/200µl)

**7. Page 10, Item 01, Semi automated coagulometer, Point No. 14, Sub Point No. c:**

**For:**

Reagent Adaptors: 1 Piece of each possible size compatible with this machine

**Read**

Deleted

**8. Page 10, Item 01, Semi automated coagulometer, Point No. 14, Sub Point No. d:**

**For:**

Stirring magnets: 4 Pieces

**Read**

Dispenser for magnetic bead/stirrer: 1 Piece.

**9. Page 10, Item 01, Semi automated coagulometer, Point No. 14, Sub Point No. e:**

**For:**

Main cable -1 piece

**Read**

Deleted

**10. Page 10, Item 01, Semi automated coagulometer, Point No. 14, Sub Point No. f:**

**For:**

Reagent container: 200 Pieces

**Read**

Deleted

**11. Page 10, Item 01, Semi automated coagulometer, Point No. 14, Sub Point No. g:**

**For:**

Reagent tubes: 200 Pieces

**Read**

Deleted

**12. Page 10, Item 01, Semi automated coagulometer, Point No. 14, Sub Point No. h:**

**For:**

Thermal Printer: 1 Piece

**Read**

Printer: 1 Piece (not needed additional if in built printer already present).

**13. Page 10, Item 01, Semi automated coagulometer, Point No. 14, Sub Point No. j:**

**For:**

Printer Cable, 1 Pc.

**Read**

Data transfer Cable, 1 Pc.

**Administrative Officer  
AIIMS, Jodhpur**