



**ALL INDIA INSTITUTE OF MEDICAL SCIENCES,  
JODHPUR**

\*\*\*\*\*

Dated: 18<sup>th</sup> February, 2016

**CORRIGENDUM**

**Rate Contract  
for  
General Hospital Items**

<b>Tender No.</b>	:	<b>Admn/RC/11/2015-AIIMS.JDH</b>
<b>NIT Issue Date</b>	:	<b>30<sup>th</sup> December, 2015</b>
<b>Last Date of Submission</b>	:	<b>15<sup>th</sup> February, 2016 at 03:00 pm</b>
<b>Extended Date</b>	:	<b>01<sup>st</sup> March, 2016 at 03:00 pm</b>
<b>Revised Last Date of Submission</b>	:	<b>04<sup>th</sup> March, 2016 at 04:00 pm</b>

Modifications for Rate Contract for supply of General Hospital Items are as under:

<b>S. No.</b>	<b>Tender Ref. No.</b>	<b>Existing Detail</b>	<b>Modifications</b>
1.	1.	<b>Item Name:</b> Adhesive Crepe Bandage  <b>Specification</b> 10 cm	<b>Item Name:</b> Latex free adhesive crepe bandage
2.	2.	<b>Item Name:</b> Adhesive Crepe Bandage  <b>Specification</b> 8 cm	<b>Item Name:</b> Latex free adhesive crepe bandage
3.	4.	<b>Item Name:</b> Transparent, Porous, Plastic, Surgical Adhesive Tape  <b>Specification</b> 1.5"	<b>Specification</b> 0.5"
4.	8.	<b>Item Name:</b> Hydrocolloid Dressing  <b>Specification</b>	<b>Specification</b> All sizes, quote separately
5.	9.	<b>Item Name:</b> Hydrofiber Dressing with Silver	<b>Specification</b> All sizes, quote separately



# ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

\*\*\*\*\*

S. No.	Tender Ref. No.	Existing Detail	Modifications
		<b>Specification</b>	
4.	11.	<b>Item Name:</b> Paper Based Adhesive Rolls 4 inch  <b>Specification</b> 9 mtr  <b>Companies</b> 3M, Suture India, Sterimed	<b>Item Name:</b> Paper Based Adhesive Rolls 3 Inch  <b>Specification</b> 9 mtr or more  <b>Companies</b> 3M, Suture India, Sterimed, Datt Medi Product
5.	12.	<b>Item Name:</b> Latex Free Silk - Like Adhesive Tape for Endotracheal Tube Fixation  <b>Specification</b>	<b>Specification</b> All Sizes, quote separately  <b>Companies:</b> 3M, Sterimed, Suture India
6.	31.	<b>Item Name</b> Abdominal Sponge	<b>Item Name</b> Abdominal sponge with X-Ray detachable thread.
7.	57.	<b>Item Name</b> Catheter Mount	<b>Companies:</b> Intersurgical, BD, Fisher & Paykel, <b>Polymed</b>
8.	69.	<b>Item Name</b> Ophthalmic Visco Elastic Device  <b>Specification</b> Sodium Hyaluronate (pfs)	<b>Specification</b> Sodium Hyaluronate (pfs), Product should be USFDA / European CE approved.
9.	70.	<b>Item Name</b> Ophthalmic Visco Elastic Device  <b>Specification</b> Hydroxy Prophyl Methyl Cellulose 2% (pfs)	<b>Specification</b> Hydroxy Prophyl Methyl Cellulose 2% (pfs), Product should be USFDA / European CE approved.
10.	71.	<b>Item Name</b> Hydrophobic Aspheric Acrylic IOL (-3 to +30 D):  <b>Specification</b> Should be single piece Hydrophobic Acrylic Aspheric Optic with square edge design Haptic should have open loop L/C shaped design Should have UV absorbers, yellow chromophore Refractive index should be more than 1.45 Should have an overall size of 13 mm and optic size 6 mm Cartridge and injector system should be capable of impacting the IOL through 2.2 mm incision Should be US FDA & European CE certified.	<b>Specification</b> Should be single piece Hydrophobic Acrylic Aspheric Optic with square edge design Haptic should have open loop L/C shaped design Should have UV absorbers, yellow chromophore Refractive index should be more than 1.45 Should have an overall size of 13 mm and optic size 6 mm Cartridge and injector system should be capable of impacting the IOL through 2.2 mm incision Should be US FDA / European CE certified.
11.	72.	<b>Item Name</b>	<b>Specification</b> Should be single piece Hydrophobic Acrylic Aspheric Optic with square edge design Haptic should have open loop L/C



# ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

\*\*\*\*\*

S. No.	Tender Ref. No.	Existing Detail	Modifications
		Hydrophobic Aspheric Acrylic IOL (-3 to +30 D):  <b>Specification</b> Should be single piece Hydrophobic Acrylic Aspheric Optic with square edge design Haptic should have open loop L/C shaped design Refractive index should be more than 1.45 Should have an overall size of 13 mm and optic size 6 mm Cartridge and injector system should be capable of impacting the IOL through 2.2 mm incision Should be US FDA & European CE certified	shaped design Refractive index should be more than 1.45 Should have an overall size of 13 mm and optic size 6 mm Cartridge and injector system should be capable of impacting the IOL through 2.2 mm incision Should be US FDA / European CE certified
7.	89.	<b>Item Name</b> Transfer blood bag adult (300 ml)	<b>Companies:</b> Terumo Penpol, Fresenius Kabi, <b>Polymed</b>
8.	90.	<b>Item Name</b> Transfer blood bag pediatric (100 ml)	<b>Companies:</b> Terumo Penpol, Fresenius Kabi, <b>Polymed</b>
9.	111.	<b>Item Name:</b> Glutaraldehyde 2.45% Solution for Sterilization  <b>Companies:</b> JnJ, 3M	<b>Companies:</b> JnJ, 3M, <b>Microgene</b>

**Page No.: 27 > Wound care management dressing**

S. No.	Tender Ref. No.	Existing Detail	Modifications
1.	1.	<b>Item Name:</b> Wound care management dressing	<b>Specifications:</b> Nano crystalline silver anti-microbial barrier dressing  <b>Sizes:</b> All sizes, quote separately
2.	2.		
3.	3.		
4.	4.		
5.	5.		
6.	6.		

**Added:**

S. No.	Tender Ref. No.	Item Name	Specification	Recommended Companies
1.	115	Double J Stent 3F	12, 14, 16, 18 cm, quote separately	Boston, Devon, Cook
2.	116	Double J Stent 4F	12, 14, 16, 18 cm, quote separately	Boston, Devon, Cook
3.	117	Double J Stent 5F	12, 14, 16, 18 cm, quote separately	Boston, Devon, Cook
4.	118	Double J Guided Wire		Boston, Devon, Cook
5.	119	Double J Stent Pusher		Boston, Devon, Cook
6.	120	Classic Surgical Caps		All can apply
7.	121	Double Lumen Polyurethane Dialysis Catheter kit	8 Fr. x Length 11 cm, Straight & Jugular 8.5 Fr. x Length 11 cm, Straight & Jugular	Covidien, Medcomp, Vygon, Newtech



**ALL INDIA INSTITUTE OF MEDICAL SCIENCES,  
JODHPUR**

\*\*\*\*\*

S. No.	Tender Ref. No.	Item Name	Specification	Recommended Companies
		(The kit must contains Polyurethane Catheter , Guide Wire , Vessel Dialators ,Introducer needle , Material Sheet, Injections caps)	11.5 Fr. * Length 15 cm, Straight & Jugular 12 Fr. * Length 13 cm., 15 cm., 20 cm, Straight & Jugular	
8.	122	Triple Lumen Polyurethane Dialysis Catheter kit (The kit must contains Polyurethane Catheter , Guide Wire , Vessel Dialators, Introducer needle , Material Sheet, Injections caps)	12 Fr. * 13 cm , 15cm , 20 cm Straight Jugular	Covidien , Medcomp, Vygon, Newtech
9.	123	Laparoscopic Ventral Hernia Repair	Polypropylene Mesh with Hydrogel anti adhesive barrier	JnJ, Bard, Covidien
10.	124	Open Umbilical Hernia Repair	Polypropylene mesh with Hydrogel film as adhesion barrier	JnJ, Bard, Covidien
11.	125	Co-axial set for biopsy	14 G, All Sizes, quote separately	Bard, Cook, Medtronic
12.	126	Co-axial set for biopsy	16 G, All Sizes, quote separately	Bard, Cook, Medtronic

**Administrative Officer**