



Date: - 28th June, 2018

Corrigendum
For
Tender for
Radiant Warmer Cum Incubator for the Department of
Pediatrics

NIT Issue Date	: 07 th May, 2018
NIT No.	: Admn/Tender/42/2018-AIIMS.JDH
Pre-Bid Meeting	: 21 st May, 2018 at 03:30 PM
Earlier Last Date of Submission	: 03 rd July, 2018 at 03:00 PM
Extended Last Date of Submission	: 18 th July, 2018 at 03:00 PM
Bid opening	: 19 th July, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 12, Point no. 25c :

For

Reusable Skin Probe – 10 Nos

Read

Reusable Skin Probe – 10 Nos OR Disposable skin probes -50 Nos

2. Page No. 11, point no. 9a :

For

Should allow tilt for Trendelenburg as well as reverse Trendelenburg position

Read

Should allow tilt for Trendelenburg as well as reverse Trendelenburg position. It would be good to have knobs outside the incubator so that these can be accessed without opening the access doors.