



Date: - 28th June, 2018

Corrigendum
For
ICU Beds for Morbid Obese Patient with Weighing
Scale for the Department of Anaesthesia

NIT Issue Date	: 07 th May, 2018
NIT No.	: Admn/Tender/46/2018-AIIMS.JDH
Pre-Bid Meeting	: 22 nd May, 2018 at 03:15 PM
Last Date of Submission	: 21 st June, 2018 at 03:00 PM
Last Date of Submission	: 09 th July, 2018 at 03:00 PM
Bid opening	: 10 th July, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 11, Point no. 2 :

For

Straight Side rails with fold away movement with illuminated hand held control panel for patient & care giver.

Read

Straight/Split Side rails with fold away movement with illuminated hand held control panel for patient & care giver.

2. Page No. 11, point no. 3 :

For

Hi-Low, head (70 degree), knee break (34 degree), TR/RTR, Shock position electrical functions

Read

Hi-Low, head (70^o+/- 10%), knee break (30^o+/- 10%), TR/RTR, Shock position electrical functions..

3. Page No. 11, point no. 4 :

For

Straight metallic collapsible side rails

Read

Straight collapsible side rails

4. Page No. 11, point no. 7 :

For

Low height of 28 cm & Max height of 80 cm

Read

Low height of 30 cm (+/-10%) & Max height of 80 cm (+/-10%)

5. Page No. 11, point no. 8 :

For

Dimensions L x W – 220 x110 cm

Read

Dimensions L x W – 220 x110 cm (+/-10%).

6. Page No. 11, point no. 14 :

For

14° Trend and 14° reverse trend for emergency situations.

Read

Minimum 12° Trend and 12° reverse trend for emergency situations

7. Page No. 11, point no. 18 :

For

Safe working load of 320 Kg.

Read

Safe working load of at least 250 Kg.