



**All India Institute of Medical Sciences
Jodhpur**

Date: - 16/07/2015

**Corrigendum
for
Delivery Beds**

NIT Issue Date	:	April 30 th , 2015.
NIT No.	:	Admn/Tender/04/2015-AIIMS.JDH
Pre Bid Meeting held on	:	May 11 th , 2015 at 03:00 PM
Last Date of Submission	:	May 22 nd , 2015 at 03:00 PM
Revised Lase Date of Submission	:	July 30 th , 2015 at 03:00 PM

1. The following revised and additional specification will be added:-

1. Page No. 10, Annexure-I, Point No. 01:

For

Overall size 2100 – 2400 mm L when extended, 900mm– 1100 mm W (with and without side rails up), 440 mm to 900 mm H (without mattress).

Read

Overall size 2100 – 2400 mm L when extended, 900mm– 1100 mm W (with and without side rails up), minimum height 440mm to 550mm and maximum height 800mm to 900 mm H (without mattress).

2. Page No. 10, Annexure-I, Point No. 03:

For

It should have control devise for making height and back adjustment [manual as well as remote control].

Read

It should have control devise for making height and back adjustments 60 degree or more [by remote control].

3. Page No. 10, Annexure-I, Point No. 06:

For

It should have three sectional mattress and seat section should have large perineal cut.

Read

It should have three sectional mattress and seat section should have large perineal cut. Should have foot support for chair position.

4. Page No. 10, Annexure-I, Point No. 15:

For

It should be able to give trendelenburg, reverse trendelenburg and 60 - 70 degree sitting position both mechanically and electronically.

Read

It should be able to give trendelenburg, reverse trendelburg 12 degree, and 60 - 70 degree sitting position.

5. Page No. 10, Annexure-I, Point No.16, sub para no. 02 :

For

Shall meet IEC-60601-1-2:2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility or should comply with 89/366/EEC; EMC-directive.

Read

Shall meet IEC-60601-2- 52 (Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility or should comply with 93/42/EEC.

6. Page No. 11, Annexure-I, Point No. 16, sub para no. 07:

For

Should be FDA, CE, UL or BIS approved product.

Read

Should be FDA or European CE approved product.

7. Page No. 11, Annexure-I, Point No. 16, sub para no. 11:

For

Electrical safety conforms to standards for Electrical Safety IEC 60601-2-38 Particular safety requirements for Electrically operated hospital beds.

Read

Electrical safety conforms to standards for Electrical Safety IEC 60601-2-38 Particular safety requirements for Electrically operated hospital beds.

OR

MDD 93/42/EEC with amendment as per 2007/47/EC

EN 60601-1-2006(3rd edition) –General safety standards

Symbol IEC 60417-5840, EN 60601-1(3rd edition): general safety requirements

IEC 60529, EN 60601-1(3rd edition): general safety requirements.

**Administrative Officer
AIIMS Jodhpur**