



Date: - 29th June 2016

Corrigendum
for
Syringe Infusion Pump

NIT Issue Date	:	01 st June 2016.
NIT No.	:	Admn/Tender/14/2016-AIIMS.JDH
Pre Bid Meeting held on	:	16 th June, 2016 at 03:45 PM
Last Date of Submission	:	23 rd June, 2016 at 03:00 PM
Revised Last Date of Submission	:	12 th July, 2016 at 03:00 PM

1. The following revised and additional specification will be added:-

1. Page No. 10, Technical Specification, Point No. 03:

For

Manufacturer should be ISO and CE certified for quality standards.

Read

Manufacturer should be ISO and European CE certified for quality standards.

2. Page No. 10, Technical Specification, Point No. 05:

For

Bolus rate should be programmable to 40 to 1000 ml/hr or more with infused volume display and one key press bolus. Reminder audio after every 0.5 ml delivered bolus.

Read

Bolus rate should be programmable to 40 to 1000 ml/hr or more with infused volume display and one key press bolus, auto bolus as a advance feature with programmable bolous rate.

3. Page No. 10, Technical Specification, Point No. 15:

For

Mounting device/ Docking Station for two or four pumps as per requirement so as to enable to power up to 2-4 pumps with one power cord when mounted on IV pole.

Read

Docking Station for two or four pumps as per requirement so as to enable to power up to 2-4 pumps with one power cord when mounted on IV pole.

Administrative Officer
AIIMS, Jodhpur