



Date: - 25th October 2016

Corrigendum
for
Equipments required for Department of Obstetrics & Gynecology

NIT Issue Date	:	21 st July 2016.
NIT No.	:	Admn/Tender/OBG/2016-AIIMS.JDH
Pre Bid Meeting held on	:	08 th August, 2016 at 03:00 PM
Extended Last Date of Submission	:	07 th November, 2016 at 03:00 PM
Revised Last Date of Submission	:	15 th November, 2016 at 03:00 PM

1. The following revised and additional specification will be added:-

1. Corrigendum, Page No. 07, Point No. 26, In High Definition camera system for Endoscopic Surgery with advance recording and data management facility, Toggle no. 03: For

- System should have facility to offer various visualization modes for surgery and recognition of the finest tissue structures and their differentiation and intraoperative perfusion assessment of tissues by fluorescence imaging.

Read

- System should have facility to offer various visualization modes for surgery and recognition of the finest tissue structures and their differentiation and intraoperative assessment to tissues by shifting different color spectrum imaging.

2. Corrigendum, Page No. 09, Point No. 36, In High Definition camera system for Endoscopic Surgery with advance recording and data management facility, In heading, LED Light Source Qty-, Toggle no. 05: For

More than 500 Hours bulb life (five extra bulbs to be supplied)

Read

Xenon Light source 300 watts or more

Xenon lamp life guarantee of 500 hours (five extra bulbs to be supplied)

Administrative Officer
AIIMS, Jodhpur