



Date: - 18th February 2016

Corrigendum
for
Equipments required for Department of
Pediatric Surgery

NIT Issue Date	:	15 th January 2016.
NIT No.	:	Admn/Tender/Pediatric Surgery/2015-AIIMS.JDH
Pre Bid Meeting held on	:	04 th February 2016 at 03:00 PM
Last Date of Submission	:	22 nd February 2016 at 03:00 PM
Revised Last Date of Submission	:	02 nd March 2016 at 03:00 PM

1. The following revised and additional specification will be added:-

- 1. Page 14, Item 04, Composite Offset Pediatric Operating Cyto-urethroscope, Point No. 1, Sub Point No. 3:**
For:
With 45 degree Angled eyepiece/ocular
Read:
45⁰ Angled ocular with 5⁰ -10⁰ Angle of view.
- 2. Page 14, Item 04, Composite Offset Pediatric Operating Cyto-urethroscope, Point No. 1, Sub Point No. 4:**
For:
Length – 150-170mm
Read:
Length – 130-170mm.
- 3. Page 14, Item 04, Composite Offset Pediatric Operating Cyto-urethroscope, Point No. 2, Sub Point No. 3:**
For:
With 45 degree Angled eyepiece/ocular
Read:
With 45 degree Angled eyepiece/ocular with 5⁰ -10⁰ Angle of view.
- 4. Page 14, Item 04, Composite Offset Pediatric Operating Cyto-urethroscope, Point No. 2, Sub Point No. 4:**
For:
Length – 210-230mm

Read:

Length – 130-170mm.

5. Page 15, Item 05a), Pediatric Retractor System-1, Point No. g, Sub Point No. II:

For:

Support Arm – 1

Read:

Deleted

6. Page 15, Item 05a), Pediatric Retractor System-1, Point No. g, Sub Point No. III:

For:

Small Curved (Wishbone) frame arm retractor-2

Read:

Small Curved (Wishbone) frame arm retractor-01

7. Page 19, Item 05c), Pediatric Open Surgical Instruments, After Point No. 22:

Add Point No. 23:

23. Stille Chirel

- 1) 25-2 8mm length
- 2) Head 28-35mm
- 3) Weight- 300-350gm
- 4) USFDA/ European CE

8. Page 19, Item 05c), Pediatric Open Surgical Instruments, After Point No. 23:

Add Point No. 24:

24. Stille Chirel

- 1) Straight
- 2) 20cm length
- 3) Tip Size
- 4) USFDA/ European CE

9. Page 19, Item 05c), Pediatric Open Surgical Instruments, After Point No. 24:

Add Point No. 25:

25. Fried man bone ronguer

- 1) 12.5 cm
- 2) Sharp tip
- 3) Straight
- 4) USFDA/ European CE

10. Page 19, Item 05c), Pediatric Open Surgical Instruments, After Point No. 25:

Add Point No. 26:

26. Bone Cutter- small straight

- 1) Stainless steel
- 2) USFDA/ European CE

11. Page 19, Item 07, Patient Warming System:

For

Control unit should be capable of warming minimum four segments at a time.

Read

Control unit should be capable of warming upto four segments at a time.

**Administrative Officer
AIIMS, Jodhpur**