



All India Institute of Medical Sciences Jodhpur

Admn/Prop/16/2021-AIIMS.JDH

Dated: - May 21, 2021

Subject: Purchase of Nervecheck Master (with all attachments and accessories) for the Intramural Research Project at AIIMS, Jodhpur on proprietary basis - **Inviting comments thereon.**

The Institute is in the purchase of Nervecheck Master (with all attachments and accessories) for the Intramural Research Project at AIIMS, Jodhpur from M/s Phi Med Europe S.L., Barcelona, Spain on proprietary basis. The proposal submitted by M/s Phi Med Europe S.L., Barcelona, Spain and PAC certification by user are attached.

The above documents are being uploaded for open information to submit objection, comments, if any from any manufacturer regarding proprietary nature of the equipment within 21days of issue giving reference Admn/Prop/16/2021-AIIMS.JDH. The comments should be received by office of Dean (Research) at AIIMS, Jodhpur on or before 11th June 2021 upto 03:00 PM failing which it will be presumed that any other vendor is having no comment to offer and case will be decided on merits.

Dean (Research)

Enclosed: Related documents enclosed.

Specifications of Equipment for quantitative sensory testing

Stimulation Methods Level & Limits

Compatible Operating Systems: Windows 10

Screen: LCD Alphanumeric

Interface between Device and PC: USB

Software to Download data: Should be included

Batteries Unit A / Unit B: Lithium-polymer 392mAh /Lithium polymer 5000mAh

Vibration Stimulator: Operating frequency (Hz):248 Hz \pm 1%

Thermal Stimulator: Temperature Range ($^{\circ}$ C):

Cold Test: 15 $^{\circ}$ C - 25 $^{\circ}$ C \pm 0.5 $^{\circ}$ C

Warm Test: 40 $^{\circ}$ C - 46 $^{\circ}$ C \pm 0.5 $^{\circ}$ C

Heat Pain Test: 32 $^{\circ}$ C - 49.5 $^{\circ}$ C \pm 0.5 $^{\circ}$ C

Time test performance (seconds)

Vibration Test: 02s \pm 1%

Cold Test: Max. 40s \pm 1%

Warm Test: Max. 40s \pm 1%

Heat Pain Test/ Warm Test by Limit: Max. 16 sec \pm 1%

Operating conditions: Temperature: 18 $^{\circ}$ C (64 $^{\circ}$ F) / 30 $^{\circ}$ C (86 $^{\circ}$ F)

Operating conditions: Humidity: 20 - 80 %

Storage conditions: Temperature: 0 $^{\circ}$ C (32 $^{\circ}$ F) / 45 $^{\circ}$ C (113 $^{\circ}$ F)

Storage conditions: Humidity: 10 - 95 %

Voltage: 100 – 240V AC

Frequency: 50 - 60 Hz

Power adapter output: 18 W





Barcelona, March 22nd of 2021

From my highest consideration;

Through this document, it is stated that the company Phi Med Europe S.L. is the owner of NerveCheck Master instruments and consumables patent as well as the only manufacturer.-

Without any other particular greetings cordially,

A handwritten signature in black ink, appearing to be "Samantha Odriozola", written in a cursive style.

A handwritten signature in blue ink, appearing to be "hmr", written in a cursive style.

Ms. Samantha Odriozola
Directora de Marketing y Ventas
Phi Med Europe S.L.