Tender

For

Equipments required for Department of Microbiology

At

All India Institute of Medical Sciences, Jodhpur

: 21st January, 2016 **NIT Issue Date**

: Admn/Tender/Microbiology/2015-AIIMS.JDH NIT No.

: 08th February, 2016 at 03:00 PM. **Pre-Bid Meeting** : 24th February, 2016 at 03:00 PM. **Last Date of Submission**



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur, Rajasthan-342005.

Telephone: 0291-2012978, email: aoadmin@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

All India Institute of Medical Sciences (AIIMS), Jodhpur, Rajasthan, an apex healthcare institute established by an Act of Parliament of India under aegis of Ministry of Health & Family Welfare, Government of India, invites sealed tenders for supply & installation of the following items for the Department of Microbiology at the institute. You are requested to quote your best offer along with the complete details of specifications, terms & conditions.

Chapter-I

S.No	NIT No.	Item Description	Qty	EMD (Rs.)
01.	Admn/Tender/Microbiology/1/2015-AIIMS.JDH	Table Top Centrifuge	4	4,000
02.	Admn/Tender/Microbiology/2/2015-AIIMS.JDH	Horizontal Autoclave	2	80,000
03.	Admn/Tender/Microbiology/3/2015-AIIMS.JDH	Autoclave (Vertical)	2	40,000
04.	Admn/Tender/Microbiology/4/2015-AIIMS.JDH	Multichannel Pipettes	4	6,400
05.	Admn/Tender/Microbiology/5/2015-AIIMS.JDH	Table Top Refrigerated Centrifuge	1	16,000
06.	Admn/Tender/Microbiology/6/2015-AIIMS.JDH	Faculty Microscopes	2	12,000
07.	Admn/Tender/Microbiology/7/2015-AIIMS.JDH	Vortex Mixer	10	2,000
08.	Admn/Tender/Microbiology/8/2015-AIIMS.JDH	Water Bath	2	2,000
09.	Admn/Tender/Microbiology/9/2015-AIIMS.JDH	Bead Bath	2	2,000
10.	Admn/Tender/Microbiology/10/2015-AIIMS.JDH	Automated Air Sampler	1	3,000

Quotation should be sealed and super-scribed with tender number and address to:

"Administrative Officer
All India Institute of Medical Sciences, Jodhpur
Basni, Phase-II
Jodhpur-342005, Rajasthan".

The sealed quotations should reach the Institute, latest by 24th February, 2016 at 03:00 PM and it will be opened on same day at 04:30 PM in the Conference Hall, Academic Block, AIIMS Jodhpur in the presence of the bidder(s) or their authorized representative(s), who will present at the scheduled date and time.

Terms & Conditions:

- 1. Preparation and Submission of Tender: The tender should be submitted in two parts i.e. Technical Bid and Financial Bid. The Technical Bid and the Financial Bid should be sealed by the bidder in two separate covers "Technical Bid for Tender for Supply of (Item Name)" and "Financial Bid for Tender for Supply of (Item Name)". Both Sealed Envelopes (separate for each NIT) should be kept in a main/ bigger envelope super-scribed as "Tender for Supply of (Item Name)".
- 2. Earnest Money Deposit: Earnest Money Deposit: The bidder shall be required to submit the Earnest Money Deposit (EMD) for an amount as per Chapter-I by way of demand drafts and Bank Guarantee only. The demand drafts shall be drawn in favour of "All India Institute of Medical Sciences, Jodhpur". The demand drafts and Bank Guarantee for earnest money deposit must be enclosed in the envelope containing the technical bid. The EMD of the successful bidder shall be returned after the successful submission of Bank Guarantee/ Security Deposit and for unsuccessful bidder(s) it would be returned after award of the contract. Bid(s) received without demand drafts of EMD will be rejected.

The firms who are registered with National Small Industries Corporation (NSIC) / OR Small Scale Industrial (SSI) are exempted to submit the EMD (copy of registration must be provide along with).

- 3. Tender Fee: Tender fee will be Non-refundable amount of Rs. 1,000/- (Rupees One Thousand only) for Each Item by way of demand drafts only. The demand drafts shall be drawn in favour of "<u>All India Institute of Medical Sciences, Jodhpur</u>". Bids received without demand draft of tender will be rejected.
- **4.** Validity: The quoted rates must be valid for a period for 180 days from the date of closing of the tender. The overall offer for the assignment and bidder(s) quoted price shall remain unchanged during the period of validity. If the bidder quoted the validity shorter than the required period, the same will be treated as unresponsive and it may be rejected.
- 5. "PRE -BID Meeting" with the intending bidders shall be held on 08th February, 2016 from 03:00 P.M. onwards at AIIMS, Jodhpur. All the prospective bidders are requested to send comments/ representations on or before pre-bid meeting. Intending bidder will be allowed to seek clarification on specification, Conditions of Contract, etc. in writing to AIIMS, Jodhpur, within 48 hours after the pre-bid meeting.
- **6.** In case the tenderer withdraws, modifies or change his offer during the validity period, bid is liable to be rejected and the earnest money deposit shall be forfeited without assigning any reason thereof. The tenderer should also be ready to extend the validity, if required, without changing any terms, conditions etc. of their original tender.

7. Delivery and Installation:

i) For goods supplied from India:

All the goods ordered shall be delivered and Installed at AIIMS, Jodhpur within **30 days** from the date of issue of supply order.

ii) For goods imported directly from abroad:

All the goods ordered shall be delivered and Installed at AIIMS, Jodhpur within **60 days** from the date of opening of Letter of Credit for shipment.

All the aspects of safe delivery, installation and commissioning shall be the exclusive responsibility of the supplier. If the supplier fails to deliver, install and commission the goods on or before the stipulated date, then a penalty at the rate of 0.5% per week of the total order value shall be levied subject to maximum of 10% of the total order value. The successful tenderer will also provide required training for supplied items at AIIMS Jodhpur. The goods should be manufactured after adoption of latest technology.

If at any time during the currency of the contract, the supplier encounters conditions hindering timely of the goods and performance of services, the supplier shall promptly inform the AIIMS, Jodhpur for extension of the delivery schedule accordingly. On receiving the supplier's communication, the AIIMS, Jodhpur shall examine the situation as soon as possible and, at its discretion, may agree to extend the delivery schedule, with or without liquidated damages for completion of supplier's contractual obligations by issuing an amendment to the contract.

In the case of package supply where the delayed portion of supply materially hampers installation and commissioning of the systems, liquidated damages charges shall be levied as above on the total value of the concerned package of the purchase order. Quantum of liquidated damages assessed and levied by the purchaser shall be final and not challengeable by the supplier.

8. Signing the Contract: - The successful bidder shall be required to execute the Contract Agreement accepting all terms and conditions stipulated herein on a non-judicial stamp paper of Rs. 100/- (Rs. One Hundred only) along with performance security within fifteen days of the issue of the Letter of notification of award. In the event of failure on the part of the successful bidder to sign the Contract

within the period stipulated above, the EMD shall be forfeited and the acceptance of BID shall be considered as cancelled.

- 9. **Performance Security:** As a guarantee towards due performance and compliance of the contract work, the successful bidder (contractor) will deposit an amount equal to 10% of order value and should be kept valid for a period of 60 day beyond completion of all the contractual obligation, including CMC period towards security deposit by way of demand draft/ bank Guarantee in favour of "All India Institute of Medical Sciences, Jodhpur "drawn on any Nationalized Bank/Scheduled Bank and payable at Jodhpur within fifteen days of the issue of the Letter of notification of award along with non-judicial stamp paper of Rs. 100/- (Contract agreement).
- 10. Incidental Services: The supplier shall be required to perform the following services:
 - a. Installation & Commissioning, Supervision and Demonstration of the goods.
 - b. Providing required jigs and tools for assembly, minor civil works required for the completion of the installation.
 - c. On Site Training to Doctors/ Technicians/ Staff is to be provided by Supplier for operation and maintenance of the equipment for a period of 30 working days after successful installation of the machine, as per direction of user department.
 - d. Supplying required number of operation & maintenance manual for the goods.
 - e. To provide non-locked open software and standard interface inter-operability conditions for networked equipment's in hospital management information system, wherever applicable.
- 11. Accessories & Consumables: The separate price list of all accessories and consumables, if any, must be attached/ enclosed along with the Financial Bid.
- 12. After Sales Service: After sales service centre should be available on 24 (hrs.) X 7 (days) X 365 (days) basis. Complaints should be attended properly, maximum within 24 hrs to ensure an uptime of minimum 95%, wherever applicable, failing which the necessary penalty measures shall be enforced.

13. Inspection:

- a. AIIMS, Jodhpur shall have the right to inspect and/or to test the goods to confirm their conformity to the NIT Specifications at no extra cost to the Purchaser.
- b. AIIMS, Jodhpur right to inspect, test and, where necessary, reject the Goods after the goods arrival at the final destination shall in no way be limited or waived by reason of the Goods having previously been inspected, tested and passed by AIIMS, Jodhpur prior to the goods shipment.
- c. The Director, AIIMS Jodhpur shall be the final authority to reject full or any part of the supply which is not confirming to the specification and other terms and conditions.
- d. No payment shall be made for rejected Stores. Rejected items must be removed by the Bidders within two weeks of the date of rejection at their own cost and replaced immediately. In case these are not removed, these will be auctioned at the risk and responsibility of the suppliers without any further notice.

14. Documents:

- a. All pages of the Tender should be numbered and indexed.
- b. The bidder shall provide in its tender the required as well as the relevant documents like technical data, literature, drawings etc. to establish that the goods and services offered in the tender fully confirm to the goods and services specified by the purchaser in the tender documents. For this purpose the bidder shall also provide a clause-by-clause commentary on the technical specifications and other technical details incorporated by the purchaser in the tender documents to establish technical responsiveness of the goods and services offered in its tender duly indicating relevant page numbers in the product literature.
- c. The bidder shall provide a list of major Government and Private Institutions where its relevant bid item has been supplied during last one year.

- **15. Manufacturer Authorisation:** The bidder (if not original equipment manufacturer must submit Original Equipment Manufacturer authorization certificate that the tenderer is authorized for selling and maintain the equipment quoted for. Performa attached at **Annexure-III.**
- **16.** The bidders are required to submit user certificate for the relevant equipment on the letter head of the institution (Government/ Private).
- 17. The successful bidder will be required to submit order copies of the supply of the equipment in Government institutions in last 12 month for rate reasonability purpose.
- **18. Insurance:** The supplier shall make arrangements for insuring the goods against loss or damage incidental to manufacture or acquisition, transportation, storage and delivery. If the equipment's is not commissioned and handed over to AIIMS, Jodhpur within specified period, the insurance will have to be extended by the supplier at their cost till the successful installation, testing, commissioning and handing over of the goods to the AIIMS, Jodhpur.

19. Tender Currencies:

- a. The bidder supplying indigenous goods or already imported goods shall quote only in Indian Rupees. Further, imported goods to be imported and supplied by the bidder are also required to be quoted in Indian Rupees.
- b. For imported goods if supplied directly from abroad, prices shall be quoted in any freely convertible currency say US Dollar, Euro, GBP or Yen. As regards price(s) for allied services, if any, required with the goods, the same shall be quoted in Indian Rupees only, if such services are to be performed /undertaken in India.
- c. Tenders, where prices are quoted in any other way shall be treated as non -responsive and rejected.
- **20. Tender Prices:** While filling up the columns of the Financial Bid, the following aspects should be noted for compliance:

For domestic goods or goods of foreign origin located within India, the prices in the corresponding Financial Bid shall be entered separately in the following manner:

- a. The price of the goods, quoted ex-factory/ ex-showroom/ ex-warehouse/ off-the-shelf, as applicable, including all taxes and duties like sales tax, CST/ VAT, CENVAT, Custom Duty, Excise Duty etc. already paid or payable on the components and raw material used in the manufacture or assembly of the goods quoted ex-factory etc. or on the previously imported goods of foreign origin quoted ex-showroom etc.;
- b. Any sales tax or other taxes and any duties including excise duty, which will be payable on the goods in India if the contract is awarded;
- c. Charges towards Packing & Forwarding, Inland Transportation, Insurance, Loading/Unloading and other local costs incidental to delivery of the goods to their final destination as specified in the List of Requirements and Financial Bid;
- d. The price of Incidental Services, as mentioned in List of Requirements and Financial Bid;
- e. The prices of Turnkey (if any), as mentioned in List of Requirements, Technical Specification and Financial Bid; and
- f. The price of annual CMC, as mentioned in List of Requirements, Technical Specification and Financial Bid.

For goods offered from abroad, the prices in the corresponding Financial Bid shall be entered separately in the following manner:

- a. The price of goods quoted FOB port of shipment, as indicated in the List of Requirements and Financial Bid;
- b. The price of goods quoted CIF port of entry in India as indicated in the List of Requirements and Financial Bid;
- c. The price of goods quoted for delivery at AIIMS, Jodhpur as indicated in the List of Requirements, Financial Bid and Consignee List;

- d. Wherever applicable, the amount of custom duty with CDEC applicable on CIF value on the goods to be imported;
- e. The charges for Loading/Unloading, Inland transportation, Insurance and other local costs, Incidental cost to delivery of the goods from the port of entry in India to AIIMS, Jodhpur, as specified in the List of Requirements and Financial Bid;
- f. The charges for Incidental Services, as in the List of Requirements and Financial Bid;
- g. The prices of Turnkey (if any), as mentioned in List of Requirements, Technical Specification and Financial Bid; and
- h. The price of annual CMC, as mentioned in List of Requirements, Technical Specification and Financial Bid.

Additional information and instruction on Duties and Taxes: If the Bidder desires to ask for excise duty, sales tax/CST / VAT/ CENVAT, Custom Duty, Service Tax, Works Contract Tax etc. to be paid extra, the same must be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such duties and taxes and no claim for the same will be entertained later.

Excise Duty:

- a. If reimbursement of excise duty is intended as extra over the quoted prices, the supplier must specifically say so also indicating the rate, quantum and nature of the duty applicable. In the absence of any such stipulation it will be presumed that the prices quoted are firm and final and no claim on account of excise duty will be entertained after the opening of tenders.
- b. If a Bidder chooses to quote a price inclusive of excise duty and also desires to be reimbursed for variation, if any, in the excise duty during the time of supply, the Bidder must clearly mention the same and also indicate the rate and quantum of excise duty included in its price. Failure to indicate all such details in clear terms may result in rejection of that tender.
- c. Subject to sub clauses (i) & (ii) above, any change in excise duty upward/downward as a result of any statutory variation in excise duty taking place within contract terms shall be allowed to the extent of actual quantum of excise duty paid by the supplier. In case of downward revision in excise duty, the actual quantum of reduction of excise duty shall be reimbursed to the purchaser by the supplier. All such adjustments shall include all reliefs, exemptions, rebates, concession etc. if any obtained by the supplier.

<u>Sales Tax:</u> If a bidder asks for sales tax/CST / VAT/CENVAT, Service Tax and Works Contract Tax to be paid extra, the rate and nature of sales tax applicable should be shown separately. The CST / VAT/CENVAT, Service Tax and Works Contract Tax will be paid as per the rate at which it is liable to be assessed or has actually been assessed provided the transaction of sale is legally liable to sales tax/ CST / VAT/CENVAT, Service Tax and Works Contract Tax and is payable as per the terms of the contract.

Octroi Duty and Local Duties & Taxes:- Normally, goods to be supplied to Government departments against Government contracts are exempted from levy of town duty, Octroi duty, terminal tax and other levies of local bodies. However, on some occasions, the local bodies (like town body, municipal body etc.) as per their regulations allow such exemptions only on production of certificate to this effect from the concerned Government department. Keeping this in view, the supplier shall ensure that the goods to be supplied by the supplier against the contract placed by the AIIMS, Jodhpur are exempted from levy of any such duty or tax and, wherever necessary, obtain the exemption certificate from the AIIMS, Jodhpur. However, if a local body still insists upon payment of such local duties and taxes, the same should be paid by the supplier to the local body to avoid delay in supplies and possible demurrage charges and obtain a receipt for the same. The supplier should forward the receipt obtained for such payment to the AIIMS, Jodhpur to enable the AIIMS, Jodhpur reimburse the supplier and take other necessary action in the matter.

<u>Customs Duty:</u> In respect of imported goods offered from abroad, the bidder shall specify the rate as well as the total amount of customs duty payable with Custom Duty Exemption Certificate, if applicable, on the quoted goods in the Financial Bid. The bidder shall also indicate the corresponding Indian Customs Tariff Number applicable for the goods.

- a. For transportation of imported goods offered from abroad, relevant instructions as incorporated shall be followed.
- b. For insurance of goods to be supplied, relevant instructions as provided shall be followed.
- c. Unless otherwise specifically indicated in this NIT document, the terms FCA, FOB, FAS, CIF, CIP etc. for imported goods offered from abroad, shall be governed by the rules & regulations prescribed in the current edition of INCOTERMS, published by the International Chamber of Commerce, Paris.
- d. The need for indication of all such price components by the bidders, as required in this clause is for the purpose of comparison of the tenders by the purchaser and will no way restrict the AIIMS, Jodhpur right to award the contract on the selected bidder on any of the terms offered.
- 21. Indian Agent:- If a foreign bidder has engaged an agent in India in connection with its bid, the foreign bidder, in addition to indicating Indian agent's commission, if any, shall also furnish the following information:
 - a. The complete name and address of the Indian Agent and its Permanent Account Number as allotted by the Indian Income Tax authority.
 - b. The details of the services to be rendered by the agent for the subject requirement.
 - c. Details of Service outlets in India, nearest to the AIIMS, Jodhpur to render services during Warranty and CMC period.

22. Firm Price

- a. Unless otherwise specified in the NIT, prices quoted by the bidder shall remain firm and fixed during the currency of the contract and not subject to variation on any account.
- b. However, as regards taxes and duties, if any, chargeable on the goods and payable, the conditions stipulated will apply.
- 23. Conversion of tender currencies to Indian Rupees: In case the bid document permits the bidders to quote their prices in different currencies, all such quoted prices of the responsive bidders will be converted to a single currency viz., Indian Rupees for the purpose of equitable comparison and evaluation, as per the closing exchange rates established by the Reserve Bank of India for similar transactions, as on the date of 'Last Date of Submission of Tender'.

24. Payment Terms:

i) Payment for goods supplied from India:

100% payment of the total order value shall be released after the successful installation/commissioning of the ordered goods against the submission of the inspection report.

ii) Payment for Imported goods:

For imported goods payment shall be made in the following manner:

- a) On shipment: 75 % payment of the contract price shall be paid 60 days after presentation of shipping documents {goods shipped shall be paid through irrevocable, non-transferable Letter of Credit (LC) opened in favour of the supplier in a bank in his country} and upon the submission of the following documents:
 - i. Four copies of Supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
 - ii. Original and four copies of the clean, on-board Bill of Lading/Airway bill, marked freight prepaid and four copies of non-negotiable Bill of Lading/Airway bill.
 - iii. Insurance Certificate;
 - iv. Certificate of origin by the chamber of commerce of the concerned country;
 - v. Certificate of country of origin;
 - vi. Manufacture's / Supplier's warranty certificate;
 - vii. Manufacturer's own factory inspection report.
- **b)** On Acceptance: 25 % payment would be made after satisfactory installation & commissioning on issuance of Inspection certificate by the AIIMS, Jodhpur.

Note:- The supplier shall not claim any interest or any other payment under the contract.

25. Custom Clearance: For the Goods to be imported and supplied, the Institute will provide Custom Duty Exemption Certificate (CDEC) to successful bidder for availing concessional rate of duty as per prevailing Custom Tariff. In case, the bidder requires CDEC certificate, then the same should be specifically mentioned in the bid. The supplier is solely responsible for getting the material clearance from customs. Institute will provide all custom documents for custom clearance on the demand of supplier. Transportation of goods up to AIIMS, Jodhpur and its successful installation and commissioning is also the responsibility of the supplier. All charges/ expenses incurred in this process will be borne by the supplier. NO DEMURRAGE / WHARFAGE CHARGES WILL BE PAYBALE BY THE INSTITUTE UNDER ANY CIRCUMSTANCES. NO ADVANCE PAYMENT WILL BE PAYABLE FOR CUSTOM CLEARANCE/ FREIGHT/INSURANCE ETC.

26. Guarantee / Warrantee Period:

- **26.1** For the equipment value upto Rs. 5 Lakh: The Tenderers must quote for **02** years comprehensive warranty (Including all Spares, Accessories and Labour) from the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. Also the bidders are requested to submit their quote (Rates) for subsequent **03** years Comprehensive Maintenance Contract (CMC) (Including All Spares, Accessories and Labour). Failure to comply this condition will entail the rejection of the bids. The price comparison shall be taking into account on basic price and post warranty CMC.
- **26.2** For the equipment value above Rs. 5 Lakh: The Tenderers must quote for **05** years comprehensive warranty (Including all Spares, Accessories and Labour) from the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. Also the bidders are requested to submit their quote (Rates) for subsequent **05** years Comprehensive Maintenance Contract (CMC) (Including All Spares, Accessories and Labour). Failure to comply this condition will entail the rejection of the bids. The price comparison shall be taking into account on basic price and post warranty CMC.
- **27. Uptime guarantee:** The firm should provide uptime guarantee of 95%

28. Downtime penalty Clause

- a. During the comprehensive warranty period, the guarantee uptime of 95% of 365 days will be ensured. In case the down time exceeds the 5% limit penalty of extension of guaranty period by two days for each additional day of down time will be enforced. The vendor must undertake to supply all spares for optimal upkeep of the equipment for at least FIVE YEARS after handling over the unit to the Institute. If accessories / other attachment of the system are procured from the third party, then the vendor must produce cost of accessory / other attachment and the CMC from the third party separately along with the main offer and the third party will have to sign the CMC with the Institute if required.
- b. The principals or their authorized service providers are required to submit a certificate that they have satisfactory service arrangements and fully trained staff available to support the uptime guarantee.
- 29. Arbitration: If any difference arises concerning this agreement, its interpretation on payment to the made there-under, the same shall be settled out by mutual consultation and negotiation. If attempts for conciliation do not yield any result within a period of 30 days, either of the parties may make a request to the other party for submission of the dispute for decision by an Arbitral Tribunal containing Sole Arbitrator to be appointed by the Secretary, Department of Legal Affairs. Such requests shall be accompanied with a panel of names of three persons to act as the sole arbitrator. In case of such arbitrator refusing, unwilling or becoming incapable to act or his mandate having been terminated under law, another arbitrator shall be appointed in the same manner from among the panel of three persons to be submitted by the claimant. The provision of Arbitration and Conciliation Act, 1990 and the rule framed there under and in force shall be applicable to such proceedings.
- **30. Subletting of Work:** The firm shall not assign or sublet the work/job or any part of it to any other person or party without having first obtained permission in writing of AIIMS, Jodhpur, which will be

at liberty to refuse if thinks fit. The tender is not transferable. Only one tender shall be submitted by one tenderer.

- **31. Breach of Terms and Conditions:** In case of breach of any terms and conditions as mentioned above, the Competent Authority, will have the right to cancel the work order/ job without assigning any reason thereof and nothing will be payable by AIIMS, Jodhpur in that event the security deposit shall also stands forfeited.
- 32. Insolvency etc: In the event of the firm being adjudged insolvent or having a receiver appointed for it by a court or any other order under the Insolvency Act made against them or in the case of a company the passing any resolution or making of any order for winding up, whether voluntary or otherwise, or in the event of the firm failing to comply with any of the conditions herein specified AIIMS, Jodhpur shall have the power to terminate the contract without any prior notice.
- 33. Force Majeure: If, at any time during the subsistence of this contract, the performance in whole or in part by either party of any obligation under this contract is prevented or delayed by reasons of any war or hostility, act of public enemy, civil commotion, sabotage, fire, floods, explosion, epidemics, quarantine restriction, strikers lockout or act of God (hereinafter referred to as events) provided notice of happening of any such eventuality is given by party to other within 21 days from the date of occurrence thereof, neither party hall by reason of such event be entitled to terminate this contract nor shall either party have any claim for damages against other in respect of such non-performance or delay in performance, and deliveries have been so resumed or not shall be final and conclusive.

Further, that if the performance in whole or in part of any obligation under this contract is prevented or delayed by reason of any such event for a period exceeding 60 days, either party may, at least option to terminate the contract.

- **34.** Bidder shall submit a copy of the tender document and addenda thereto, if any, with each page of this document should be signed and stamped to confirm the acceptance of the entire terms & conditions as mentioned in the tender enquiry document.
- **35.** The quantity of item given in the tender is tentative, which may be increased or decreased as per the institute's requirement.
- **36.** Signed & stamped compliance sheet of the technical specification of the goods with technical printed literature must be enclosed with the bid.
- **37.** After due evaluation of the bid(s) Institute will award the contract to the lowest evaluated responsive tenderer
- **38.** Conditional bid will be treated as unresponsive and it may be rejected.
- **39. Demonstration:** AIIMS Jodhpur reserves the right to ask the tenderers for arranging demonstration of their equipment for which rates have been quoted, to the concerned committee, if required.
- **40.** The Institute reserves the right to accept in part or in full or reject any or more tender(s) without assigning any reasons or cancel the tendering process and reject all tender(s) at any time prior to award of contract, without incurring any liability, whatsoever to the affected bidder or bidder(s).

41. Applicable Law:

• The contract shall be governed by the laws and procedures established by Govt. of India, within the framework of applicable legislation and enactment made from time to time concerning such Commercial dealings / processing.

- Any disputes are subject to exclusive jurisdiction of Competent Court and Forum in Jodhpur, Rajasthan, India only.
- The Arbitration shall be held in accordance with the provisions of the Arbitration and Conciliation Act, 1996 and the venue of arbitration shall be at Jodhpur. The decision of the Arbitrator shall be final and binding on both the partied.
- Force Majeure: Any delay due to Force Majeure will not be attributable to the supplier.

Administrative Officer AIIMS, Jodhpur

Annexure-I

S.N	Item	Technical Specification	Qty
0		•	
01	Table Top Centrifuge	 Table Top Version Tube capacity: Size 5-15ml with reduction adaptors suitable for vacutainer tubes 75mm and 100 mm long Digital timer Strong fabricated & corrosion resistant steel 	4
		 5. Control panel- for start/stop switch, dynamic brakes, step less speed regulator with zero start switch& sped indicator with timer & protective fuses 6. Door interlock 7. Maintenance free brushless drive motor with exact speed pre- 	
		selection & display. Speed range: 100 to 10,000 rpm & above, accuracy 1 rpm 8. Centrifuge complete with Swig & basic rotors & four buckets- 01 set 9. Power supply: power input: 220- 240 V AC, 50Hz.	
02	Horizontal Autoclave	Stainless Steel of SS-304 grade, Triple walled with a steel jacket, separate boiler, inner and outer chamber, jacket and boiler made of stainless steel, The pressure is adjustable from 5 psi to 20 psi with automatic pressure control switch, which controls the pressure with an accuracy of +/- 1 psi., LW cut off, auto safety door lock, mounted on rigid tabular stand. Spring loaded safety valve of stainless steel. Inner dimensions of chamber: (approx.) cylindrical 600mm diameter x 1100mm deep, 9KW (approx.) electrically heated, Working pressure: 1.2 -1.5kg/sq. cm. There should be provision for vaccum for drying of racks and porous loads. Automatic Pressure Control Switch- To cut-off the current from the heating elements, when the desired/ set pressure value level is attained inside the chamber and restarts the mechanism once the pressure inside the chamber falls from the desired level. Automatic Water Cut-off Device – To protect the heaters from running dry and to ensure that the machine is automatically switched off in case the desired water level falls below the prescribed level. Temperature Indicator to indicate the temperature inside the chamber. Timer with Alarm System - To regulate the sterilization time of the media to be sterilized and when the desired time is passed, with a	2
03	Autoclave (Vertical)	 buzzer, to that the completion of sterilization cycle. Operational Requirements: Suitable for hospital dressings, linen, surgical instruments, glassware, culture media and laboratory ware etc. Capacity of 80 liters or more. Technical Specifications:- Single door high pressure steam sterilizer with double/triple walled, steam jacket and separate boiler. Material of construction: Sterilizer chamber SS 304; Door SS 304; Jacket MS; Loading carriage SS 316; Transfer trolley: MS, painted; Door Gasket: Silicon or better; Insulation: fibre glass resin bonded wool or better; Insulation cover: SS sheets. Chamber capacity as per requirement Operating temperature 121°C - 138°C pressure 1.1 to 2.2 kg/cm² of steam pressure Sterilizer should be provided with steam generator. Spring loaded safety valves and automatic vacuum breaker for jacket. 	2

9. Removable plug screen for chamber drain. 1) SS baffle for even steam distribution in the chamber. 1) Safety valve protection against poor pressure. 3) Safety lock for door: pressure lock safety device. 4) Low water off. 4. System Configuration Accessories, spares and consumables: a) System as specified b) Should provide a sufficient quality of consumable along with the equipment 5. Environmental factors: Shall meet IEC-60601-1-2: 200(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. 6. Power Supply: Power input to be 220-240VAC, 50Hz/440 V 3 Phase as appropriate and fitted with plug compatible with local sockets 1) Light weight electronic pipethe for high professional standards that provide optimal support in work Only one multi-function rocker for liquid aspiration & dispensing 3) Piston should automatically return zero position when tip is ejected 4) Spring loaded tip cone that provide maximum tightness with minimal attachment force 3) Pipette should have Li polymer battery that provide long service life without charging 6) Pipette should bave Li polymer battery that provide long service life without charging 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl-2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000
i) Safety valve protection against poor pressure. j) Safety lock for door: pressure lock safety device. k) Low water off. 4. System Configuration Accessories, spares and consumables: a) System as specified- b) Should provide a sufficient quality of consumable along with the equipment 5. Environmental factors: Shall meet IEC-60601-1-2: 200(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility: 6. Power Supply: Power input to be 220-240VAC, 50Hz,/440 V 3 Phase as appropriate and fitted with plug compatible with local sockets l) Light weight electronic pipette for high professional standards that provide optimal support in work 2) Only one multi-function rocker for liquid aspiration & dispensing 3) Piston should automatically return zero position when tip is ejected 4) Spring loaded tip cone that provide maximum tightness with minimal attachment force 5) Pipette should have Li polymer battery that provide long service life without charging 6) Pipette should work continuously while changing 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl-2 each should be supplied with 1000 disposal tips for all 11) Should have documentation certificate of calibration and inspection from factory. 2) Table Top 1) Should have documentation certificate of calibration and inspection from factory. 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro cen
j) Safety lock for door: pressure lock safety device. k) Low water off. 4. System Configuration Accessories, spares and consumables: a) System as specified- b) Should provide a sufficient quality of consumable along with the equipment 5. Environmental factors: Shall meet IEC-60601-1-2: 200(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. 6. Power Supply: Power input to be 220-240VAC. Solfa, 2440 V 3 Phase as appropriate and fitted with plug compatible with local sockets 1 Pipettes 1 Pipettes 2 Display of the multi-function rocker for liquid aspiration & dispensing Piston should automatically return zero position when tip is ejected spring loaded tip cone that provide maximum tightness with minimal attachment force 5 Pipette should water Li polymer battery that provide long service life without charging 6 Pipette should work continuously while changing 7 Parameter should be in the same position regardless of the mode 8 Provision to autoclave the lower parts 9 Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl-2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1 Table Top Refrigerated Centrifuge Table Top Programmable speed with separate short spin key (in seconds). 3 Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4 Automatic rotor recognition and automatic imbalance detection. 5 Double lid locking system for maximum safety. 6 Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7 Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8 Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
k) Low water off. 4. System Configuration Accessories, spares and consumables: a) System as specified- b) Should provide available spares and consumables for at least 10 years c) Should provide a sufficient quality of consumable along with the equipment 5. Environmental factors: Shall meet IEC-60601-1-2: 200(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. 6. Power Supply: Power input to be 220-240VAC. 50Hz/440 V 3 Phase as appropriate and fitted with plug compatible with local sockets l) Light weight electronic pipette for high professional standards that provide optimal support in work Only one multi-function rocker for liquid aspiration & dispensing 3) Piston should automatically return zero position when tip is ejected 4) Spring loaded tip cone that provide maximum tightness with minimal attachment force 5) Pipette should have Li polymer battery that provide long service life without charging 6) Pipette should work continuously while changing 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl-2 each should have adocumentation certificate of calibration and inspection from factory. 10 Table Top Refrigerated Centrifuge 10 Should have documentation certificate of calibration and inspection from factory. 11 Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 22 Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and m
4. System Configuration Accessories, spares and consumables: a) System as specified- b) Should provide a vailable spares and consumables for at least 10 years c) Should provide a sufficient quality of consumable along with the equipment 5. Environmental factors: Shall meet IEC-60601-1-2: 200(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. 6. Power Supply: Power input to be 220-240VAC. Soltz, 440 V 3 Phase as appropriate and fitted with plug compatible with local sockets 1 Pipettes 1 Pipettes 2 Only one multi-function rocker for liquid aspiration & dispensing 3 piston should automatically return zero position when tip is ejected 4 Spring loaded tip cone that provide maximum tightness with minimal attachment force 5 Pipette should work continuously while changing 7 Parameter should be in the same position regardless of the mode 8 Provision to autoclave the lower parts 9 Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10 Should have adjustable volume range from 10-100µl & 30-300µl-2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1 Table Top Refrigerated Centrifuge Table Top Refrigerated Contriled 1 Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2 Programmable speed with separate short spin key (in seconds). 3 Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4 Automatic rotor recognition and automatic imbalance detection. 5 Double lid locking system for maximum safety. 6 Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7 Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8 Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
a) System as specified- b) Should provide available spares and consumables for at least 10 years c) Should provide a sufficient quality of consumable along with the equipment 5. Environmental factors: Shall meet IEC-60601-1-2: 200(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. 6. Power Supply: Power input to be 220-240VAC, 50Hz,/440 V 3 Phase as appropriate and fitted with plug compatible with local sockets 1 Light weight electronic pipette for high professional standards that provide optimal support in work 2 Only one multi-function rocker for liquid aspiration & dispensing 3) Piston should automatically return zero position when tip is ejected 4) Spring loaded tip cone that provide maximum tightness with minimal attachment force 5) Pipette should have Li polymer battery that provide long service life without charging 6) Pipette should have Li polymer battery that provide long service life without charging 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl-2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge
a) System as specified- b) Should provide available spares and consumables for at least 10 years c) Should provide a sufficient quality of consumable along with the equipment 5. Environmental factors: Shall meet IEC-60601-1-2: 200(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. 6. Power Supply: Power input to be 220-240VAC, 50Hz,/440 V 3 Phase as appropriate and fitted with plug compatible with local sockets 1 Light weight electronic pipette for high professional standards that provide optimal support in work 2 Only one multi-function rocker for liquid aspiration & dispensing 3) Piston should automatically return zero position when tip is ejected 4) Spring loaded tip cone that provide maximum tightness with minimal attachment force 5) Pipette should have Li polymer battery that provide long service life without charging 6) Pipette should have Li polymer battery that provide long service life without charging 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl-2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge
b) Should provide available spares and consumables for at least 10 years c) Should provide a sufficient quality of consumable along with the equipment 5. Environmental factors: Shall meet IEC-60601-1-2: 200(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. 6. Power Supply: Power input to be 220-240VAC, 50Hz,/440 V 3 Phase as appropriate and fitted with plug compatible with local sockets 1 Pipettes 1 Pipettes 2 Only one multi-function rocker for liquid aspiration & dispensing 3 Piston should automatically return zero position when tip is ejected 4 Spring loaded tip cone that provide maximum tightness with minimal attachment force 5 Pipette should have Li polymer battery that provide long service life without charging 6 Pipette should work continuously while changing 7 Parameter should be in the same position regardless of the mode 8 Provision to autoclave the lower parts 9 Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10 Should have adjustable volume range from 10-100µl & 30-300µl-2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 10 Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2 Programmable speed with separate short spin key (in seconds). 3 Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4 Automatic rotor recognition and automatic imbalance detection. 5 Double lid locking system for maximum safety. 6 Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7 Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8 Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
e) Should provide a sufficient quality of consumable along with the equipment 5. Environmental factors: Shall meet IEC-60601-1-2: 200(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. 6. Power Supply: Power input to be 220-240VAC, 50Hz/440 V 3 Phase as appropriate and fitted with plug compatible with local sockets 10.4 Multichanne I Pipettes 10.5 I Table Top Refrigerated Centrifuge 10.6 Table Top Refrigerated Centrifuge 10.7 Should have documentation certificate of calibration and inspection from factory. 11. Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 12. Programmable speed with separate short spin key (in seconds). 13. Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 14. Automatic rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed. 15. Should have speed MCF.
c) Should provide a sufficient quality of consumable along with the equipment 5. Environmental factors: Shall meet IEC-60601-1-2: 200(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. 6. Power Supply: Power input to be 220-240VAC, 50Hz,440 V 3 Phase as appropriate and fitted with plug compatible with local sockets. 1) Light weight electronic pipette for high professional standards that provide optimal support in work. 2) Only one multi-function rocker for liquid aspiration & dispensing. 3) Piston should automatically return zero position when tip is ejected. 4) Spring loaded tip cone that provide maximum tightness with minimal attachment force. 5) Pipette should have Li polymer battery that provide long service life without charging. 6) Pipette should be with continuously while changing. 7) Parameter should be in the same position regardless of the mode. 8) Provision to autoclave the lower parts. 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels. 10) Should have adjustable volume range from 10-100µl & 30-300µl-2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 11) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
equipment 5. Environmental factors: Shall meet IEC-60601-1-2: 200(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. 6. Power Supply: Power input to be 220-240VAC, 50Hz,/440 V 3 Phase as appropriate and fitted with plug compatible with local sockets 1) Light weight electronic pipette for high professional standards that provide optimal support in work 2) Only one multi-function rocker for liquid aspiration & dispensing 3) Piston should automatically return zero position when tip is ejected 4) Spring loaded tip cone that provide maximum tightness with minimal attachment force 5) Pipette should have Li polymer battery that provide long service life without charging 6) Pipette should work continuously while changing 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have adjustable volume range from 10-100µl & 30-300µl-2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
5. Environmental factors: Shall meet IEC-60601-1-2: 200(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility 6. Power Supply: Power input to be 220-240VAC, 50Hz/440 V 3 Phase as appropriate and fitted with plug compatible with local sockets 1. Light weight electronic pipette for high professional standards that provide optimal support in work 2. Only one multi-function rocker for liquid aspiration & dispensing 3. Piston should automatically return zero position when tip is ejected 4. Spring loaded tip cone that provide maximum tightness with minimal attachment force 5. Pipette should have Li polymer battery that provide long service life without charging 6. Provision to autoclave the lower parts 9. Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10. Should have adjustable volume range from 10-100µl & 30-300µl-2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1. Should have adjustable volume range from 10-100µl & 30-300µl-2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 2. Programmable speed with separate short spin key (in seconds). 3. Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4. Automatic rotor recognition and automatic imbalance detection. 5. Double lid locking system for maximum safety. 6. Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7. Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8. Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
General Requirements of Safety for Electromagnetic Compatibility. 6. Power Supply: Power input to be 220-240VAC, 50Hz/440 V 3 Phase as appropriate and fitted with plug compatible with local sockets 1 Light weight electronic pipette for high professional standards that provide optimal support in work 2 Only one multi-function rocker for liquid aspiration & dispensing 3 Piston should automatically return zero position when tip is ejected 4 Spring loaded tip cone that provide maximum tightness with minimal attachment force 5 Pipette should have Li polymer battery that provide long service life without charging 6 Pripette should work continuously while changing 7 Parameter should be in the same position regardless of the mode 8 Provision to autoclave the lower parts 9 Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10 Should have adjustable volume range from 10-100µl & 30-300µl-2 each should be supplied with 1000 disposal tips for all 8 Should have documentation certificate of calibration and inspection from factory. 1 Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2 Programmable speed with separate short spin key (in seconds). 3 Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4 Automatic rotor recognition and automatic imbalance detection. 5 Double lid locking system for maximum safety. 6 Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7 Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8 Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
6. Power Supply: Power input to be 220-240VAC, 50Hz,/440 V 3 Phase as appropriate and fitted with plug compatible with local sockets 1) Light weight electronic pipette for high professional standards that provide optimal support in work 2) Only one multi-function rocker for liquid aspiration & dispensing 3) Piston should automatically return zero position when tip is ejected 4) Spring loaded tip cone that provide maximum tightness with minimal attachment force 5) Pipette should have Li polymer battery that provide long service life without charging 6) Pipette should work continuously while changing 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
Multichanne 1 Pipettes
Multichanne 1 Pipettes 1 Light weight electronic pipette for high professional standards that provide optimal support in work 2 Only one multi-function rocker for liquid aspiration & dispensing 3 Piston should automatically return zero position when tip is ejected 4 Spring loaded tip cone that provide maximum tightness with minimal attachment force 5 Pipette should have Li polymer battery that provide long service life without charging 6 Pipette should work continuously while changing 7 Parameter should be in the same position regardless of the mode 8 Provision to autoclave the lower parts 9 Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10 Should have adjustable volume range from 10-100μl & 30-300μl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1 Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2 Programmable speed with separate short spin key (in seconds). 3 Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4 Automatic rotor recognition and automatic imbalance detection. 5 Double lid locking system for maximum safety. 6 Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7 Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8 Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
optimal support in work 2) Only one multi-function rocker for liquid aspiration & dispensing 3) Piston should automatically return zero position when tip is ejected 4) Spring loaded tip cone that provide maximum tightness with minimal attachment force 5) Pipette should have Li polymer battery that provide long service life without charging 6) Pipette should work continuously while changing 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
2) Only one multi-function rocker for liquid aspiration & dispensing 3) Piston should automatically return zero position when tip is ejected 4) Spring loaded tip cone that provide maximum tightness with minimal attachment force 5) Pipette should have Li polymer battery that provide long service life without charging 6) Pipette should work continuously while changing 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
3) Piston should automatically return zero position when tip is ejected 4) Spring loaded tip cone that provide maximum tightness with minimal attachment force 5) Pipette should have Li polymer battery that provide long service life without charging 6) Pipette should work continuously while changing 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100μl & 30-300μl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
4) Spring loaded tip cone that provide maximum tightness with minimal attachment force 5) Pipette should have Li polymer battery that provide long service life without charging 6) Pipette should work continuously while changing 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl-2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
attachment force 5) Pipette should have Li polymer battery that provide long service life without charging 6) Pipette should work continuously while changing 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100μl & 30-300μl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
5) Pipette should have Li polymer battery that provide long service life without charging 6) Pipette should work continuously while changing 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
without charging 6) Pipette should work continuously while changing 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
 6) Pipette should work continuously while changing 7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100μl & 30-300μl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
7) Parameter should be in the same position regardless of the mode 8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
8) Provision to autoclave the lower parts 9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
9) Should have provision for removing individuals channels (In Multichannel Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
Pipette) to adjust the distance between channels 10) Should have adjustable volume range from 10-100µl & 30-300µl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
10) Should have adjustable volume range from 10-100μl & 30-300μl- 2 each should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
should be supplied with 1000 disposal tips for all Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
Should have documentation certificate of calibration and inspection from factory. 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
Table Top Refrigerated Centrifuge 1) Should be Microprocessor Controlled with large LC Display of time, speed and temperature. 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
Refrigerated Centrifuge of time, speed and temperature. Programmable speed with separate short spin key (in seconds). Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. Double lid locking system for maximum safety. Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
Centrifuge 2) Programmable speed with separate short spin key (in seconds). 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
 3) Brushless maintenance free motor drive with low noise levels less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
less than 63 db at Max speed. 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
 4) Automatic rotor recognition and automatic imbalance detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
detection. 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
 5) Double lid locking system for maximum safety. 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
 6) Maximum speed 15,000 to 17,000 rpm with maximum RCF 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
 20,000 to 30,000 x g with speed/RCF increment in steps of 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
 1,000. 7) Wide temperature range of -15°C to +4°C and must be able to maintain 4°C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
 7) Wide temperature range of -15⁰C to +4⁰ C and must be able to maintain 4⁰C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
maintain 4 ⁰ C at Max speed. 8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
8) Angle rotor accommodating 24 X 1.5/2.0 ml Micro centrifuge Tubes with Max speed/RCF.
Tubes with Max speed/RCF.
9) Angle Rotor accommodating 6 X 50/15 ml Falcon Tubes Max
speed up to 12,000 rpm/2,000 x g RCF.
10) Rotor and Chamber should be made of chemical resistant and
rust free material.
Suitable constant power supply backup / UPS.
Machine should be EN-61010-2-020 certified.
06FacultyOptical systemInfinitely corrected system2
Microscopes
Focus Vertical stage movement
, , , , , , , , , , , , , , , , , , , ,
25mm or more for course

			movement 1micron or less for fine stroke			
		Illuminator	Lamp house for 100 watts halogen lamp with DIC			
		Revolving nose piece	upgradable. :Reversed sextuple revolving			
			nose piece should be upgradable to DIC in future			
		Objectives	Plan achromatic 2X N.A 0.06 Plan achromatic 4X N.A 0.10 Plane achromatic 10X N.A 0.25			
			Plane achromatic 40X N.A 0.65 (spring) Plane achromatic 100X N.A 1.25 (spring & oil)			
		Observation field	Wide field trinocular eye piece tube with 10X eye pieces of 25mm or more F.O.V			
		Stage	Ceramic coated surface mechanical stage with right hand low drive controlled with left hand for two specimens.			
		Condenser	Swing out condenser usable for 2X-100X.			
		Camera & software	Digital pool CCD camera approx. 3MP/4MP, with 10 bit digitalization, 2048X1500. Exposure time 1.6ms to 17.8 m in 1 µs increments with fire wire port. Software to capture			
		Computer system	and image processing. i5 processor, 4GB RAM, 500GB HDD, DVR R/W, TFT 20".			
		multicolor 5-8 filters posi	gradable to fluorescence attachment with			
07	Vortex	• Speed range 500-3000 rpm,	with manually controllable knob	10		
	Mixer	 Acceleration time – 3 sec Orbit-2mm. 				
		Power supply: External power supply DC 12 V, 500mA				
08	Water Bath	• To be supplied with DC adapter if not built in. Temperature range from ambient temperature 0° to 100°C.				
	, attributi	Thermostatic control with an ac	curacy of plus minus 0.5°C	2		
			steel and outside mild steel sheet			
			np, temperature control knob and 20/230 volts AC supplied with or			
		without stirring arrangement wi	==			

		Lid of water	er bath is ma	de of stainle	ss steel 304 (Qlty.		
			L	M	D			
		SBS	300m	250m	175m	Suitabl	14	
		-1	m	230III m	m 173III	e for 2	Ltr	
		-1	111	111	111	racks	S	
			_		s should be running prote	there for hig	h/low	
		_		-		arrier racks s	hould	
				and test tube		differ facks s	nouid	
		• A coc	k should be	e provided	to facilitate	draining of	bath	
		content				_		
						there to pi	revent	
				formation of	_	all the acces	coriec	
		1	- 1		he equipmen		501105	
		_		•		ied or equi	valent	
			d of repute.			1		
				9001:2000	or other	equivalent q	uality	
		certific			1.0			
				-		th functionin	g e.g.	
09	Bead Bath					e equipment. without long	warm un	2
	Doug Buth	delays.	delile ve ne v	v see tempere	mare quiekry	without fong	warm ap	
			have Stainle	ess body and	lid			
				re: $+5^{\circ}$ C to 8				
					50/60Hz, 500	0W, 5.0A		
				emperature j	protection			
		,	6) With microprocessor control 7) With Digital microprocessor control					
			 7) With Digital microprocessor control 8) Temperature Uniformity +/- 0.2% at 37°C 					
		,		onth warranty				
		10) The capacity of tank should be 8-20 liters.						
		11) 4-8 liters of additional beads apart from the tank capacity should be						
10	Automated	supplie 1) Should		nfigurations	r Total Such	ended Partic	ulates	1
10	Air Sampler		PM10 and F		s. Total Susp	chaca rartic	uraces	_
		, , , ,			peed controll	ed brushels b	lower	
			_	_	and 2 filter	holders for	easy	
			ge in the fiel					
				-		ock, wide gi	_	
						er friendly sa the selection		
						heir own pro		
				nin to 168 ho			6	
		_	-		-	ntrolled syste		
						low tempera		
		ambient and venture pressures for true mass or volumetric						
		flow standardized with selectable reference temperature. 5) Measured parameters should be logged every five seconds and						
		1	-			hours run p		
				_		reports show		
						y should ab	ole to	
		downlo	paded to a PC	C via Moden	n (optional).			

- 6) Flow Range: 100 1400 L/min, **Minimum sampling Volume: 999 Litre.**
- 7) Sampling speed: approx. 100litres/min. or more
- 8) Minimum sampling volume: 999 Litres
- 9) Air Velocity is 0.45m/sec
- 10) Horizontal aspiration (Laminar), the air impact speed to the culture medium is 20m/sec.
- 11) Table speed:1rpm, correction table
- 12) Blower Motor: 1/4 hp
- 13) Flow Stability: \pm 1% of set flow rate
- 14) Power Requirements: 220-240 VAC, (**Preferred also if battery operated**) 50 Hz, 15 AMP.
- 15) Time (Selection):0 to 999sec.
- 16) Autoclavable metal top.
- 17) Collection method: Sieve impaction, Particle capture mechanism: Sieve impaction
- 18) Particle diameter cut size :1µm, Perforation diameter: 1mm
- 19) Plate Size: Standard plate (100mm), SS Cone compatible to: 90mm standard culture plate
- 20) Exchangeable battery pack remote operation: 7hrs
- 21) Mains charger: 240 Volts AC
- 22) Anodized aluminium housing, Coated Aluminum Body
- 23) Carry Bags
- 24) Norm threaded hole for camera tripod
- 25) Autoclavable perforated lid
- 26) Programmable start delay function (1 to 60mins)
- 27) Preprogrammed collection volumes; 10,20,50,100,200,250,500,750 and 100 litres
- 28) 1 volume of choice programmable
- 29) Adjustable sampling head (any angle between horizontal & vertical)
- 30) Factory calibrated (with certificate) with in house calibration facility.
- 31) Should be USFDA or CE or BIS certified.
- 32) Allowable environmental temperature operating range: 0-50°C
- 33) Detachable base: Inlet head should be easily attached on site.
- 34) All accession required for functioning and calibration to be provided.

Annexure-II

TECHNICAL BID

Name of Firm/Contractor/Supplier	
Complete Address & Telephone No.	
Name of Proprietor/Partner/Managing Director/Director.	
Phone No:-	
Mobile No:-	
Email Id:-	
Name and address of service centre nearby Jodhpur.	
Whether the firm is a registered firm Yes/No (attached copy of certificate).	
PAN No. (enclose the attested copy of PAN Card).	
Service Tax No. (enclose the attested copy of Service Tax Certificate).	
VAT No. (enclose the attested copy of VAT Certificate).	
Whether the firm has enclosed the Bank Draft/Pay Order/Banker's cheque of Earnest Money Deposit.	
Whether the Firm/Agency has signed each and every page of Tender/NIT.	
Please provide full list of consumables.	
Any other information, if necessary.	

Authorized signatory of the bidder with seal.

Annexure-III

MANUFACTURER'S / PRINCIPAL'S AUTHORIZATION FORM

10	
The Administrative Officer, All India Institute of Medical Sciences, Jodhpur	
Sir,	
TENDER:	
we,	who are
established and reputable manufacturers of	, having
factories at and	, hereby
authorize Messrs	(name and address of agents)
to bid, negotiate and conclude the contract with you	against Tender
No fo	or the above goods manufactured by
us. No company or firm or individual other than Mes	srs are
authorized to bid, negotiate and conclude the contra	ct in regard to this business
against this specific tender.	
We hereby extend our full guarantee and warranty the goods offered for supply against this tender by the	-
The authorization is valid up to	
	Yours faithfully,
	(Name) on behalf of Messrs(Name of manufacturers)/Principal.

Annexure-IV

Financial Bid

A) FINANCIAL BID FOR DOMESTIC GOODS OR GOODS OF FOREIGN ORIGIN LOCATED WITHIN INDIA OR GOODS TO BE IMPORTED AND SUPPLIED AGAINST PAYMENT IN INDIAN RUPEES

1	2	3	4		5			6			
Schedul	Brief Description of Goods	Country of Origin	Quantity (Nos.)	Ex - factory/ Ex- warehous e /Ex- showroo m /Off- the shelf (a)	Excise Duty(if any) [%age & & value] (b)	Sales Tax/ VAT(if any) [%age & value] (c)	Packing and Forwarding charges (d)	Inland Transportation, Insurance, loading/ unloading and Incidental costs at AIIMS- Jodhpur (e)	Incidental Services (including Installation & Commissioning, Supervision, Demonstration and Training) at AIIMS- Jodhpur (f)	Unit Price (at AIIMS- Jodhpur) basis (g)= (a+b+c+d+e+ f)	Total Price (at AIIMS- Jodhpur) basis (Rs.) = {4 x 5(g)}

Total Tender price in Rupees: _			
In words:			

Note: -

- 1. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail.
- 2. The charges for Annual CMC after warranty shall be quoted separately.
- 3. The Bidder must quote price for "GOODS TO BE IMPORTED AND SUPPLIED AGAINST PAYMENT IN INDIAN RUPEES" after having taken in to account, the provision of Custom Duty Exemption Certificate (CDEC) by the Purchaser, as per Customs Tariff Act.

Place:	Name:
Date:	Business Address:
	Signature of Bidder
	Seal of the Bidder:

<u>Financial Bid</u> <u>B) FINANCIAL BID FOR GOODS TO BE IMPORTED FROM ABROAD</u>

1	2	3	4		5			
						Price per unit (Rs.)		Total price on
Schedule	Brief Description of Goods	Country of Origin	Quantity (Nos.)	FOB price at port/ airport of Lading (a)	Carriage & Insurance (port of loading to port of entry) and other Incidental costs** (b)	Incidental Services (including Installation & Commissioning, Supervision, Demonstration and Training) at AIIMS-Jodhpur **	Unit Price on DDP AIIMS- Jodhpur + Extended Insurance (local transportation and storage) (d) = a+b+c	Destination + Insurance (local transportation and storage) $= \{4X \ 5 \ (d)\}$

** To be paid in Indian Currency (Rs.)	
Total Tender price in foreign currency: _	
In words:	
Note: -	

- 1. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail.
- 2. The charges for Annual CMC after warranty shall be quoted.
- 3. The Bidder will be fully responsible for the safe arrival of the goods AIIMS-Jodhpur in good condition as per terms of DDP as per INCOTERMS, if applicable.

Indian Agent:	
Indian Agency Commission% of FOB	
Place:	Name:
Date:	Business Address:
	Signature of Bidder
	Seal of the Bidder:

Financial Bid

C) FINANCIAL BID FOR ANNUAL COMPREHENSIVE MAINTENANCE CONTRACT AFTER WARRANTY PERIOD:

1. S.No.	2. DESCRIPTION OF GOODS	3. QUANTITY. (Nos.)	4. Annual Comprehensive Maintenance Contract Cost for Each Unit year wise*. 1 st 2 nd 3 rd 4 th 5 th a b c d e			5 th	5. Total Annual Comprehensive Maintenance Contract Cost for 5 Years [3 x (4a+4b+4c+4d+4e)]	

^{*} After completion of Warranty period.

Service Tax: Whether extra or inclusive, if extra, indicates the rate_____.

NOTE:-.

- 1. In case of discrepancy between unit price and total prices, THE UNIT PRICE shall prevail.
- 2. The cost of Comprehensive Maintenance Contract (CMC) which includes preventive maintenance including testing & calibration as per technical/ service /operational manual, labour and spares, after satisfactory completion of Warranty period may be quoted as per NIT conditions on yearly basis for complete equipment and Turnkey (if any).
- 3. The cost of CMC may be quoted along with taxes applicable. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
- 4. Cost of CMC will be added for Ranking/Evaluation purpose.
- 5. All software updates should be provided free of cost during CMC period.
- 6. The stipulations in Technical Specification will supersede above provisions
- 7. The supplier shall keep sufficient stock of spares required during Annual Comprehensive Maintenance Contract period. In case the spares are required to be imported, it would be the responsibility of the supplier to import and get them custom cleared and pay all necessary duties.

Date:	
Place:	Name:
	Business Address:
	Signature of Bidder:
	Seal of the Bidder: